AMERICAN ASSOCIATION OF BIOANALYSTS
205 West Levee St. Brownsville, TX 78520
(281) 436-5357 - (800) 234-5315 - Fax (713) 781-5008

PARTICIPANT STATISTICS **FIRST QUADRIMESTER 2010**

CARDIAC MARKERS / ISOENZYMES

	Specimen 1		Specimen 2		Specimen 3		Specimen 4		Specimen 5	
CK-2 (CK-MB)	Elevated	Unelevated	Elevated	Unelevated	Elevated	Unelevated	Elevated	Unelevated	Elevated	Unelevated
Biosite TriageCardiac Panel	1		1		1		1			1
LifeSign MI	1		1		1		1			1
Nanogen Cardiac STATus	5		5		5		5			5
Total Population	7	0	7	0	7	0	7	0	0	7
Flagging		*****		*****		*****		*****	*****	

	Specimen 1		Specimen 2		Specimen 3		Specimen 4		Specimen 5	
Myoglobin, qualitative	Elevated	Unelevated	Elevated	Unelevated	Elevated	Unelevated	Elevated	Unelevated	Elevated	Unelevated
LifeSign MI 668		1	1			1		1	1	
Nanogen Cardiac STATus 710		4	3	1		4		4	4	
Roche Cardiac Reader 822		1	1			1		1	1	
Total Population	0	7	6	1	1	6	0	7	7	0
Flagging	*****			*****	*****		*****			*****

	Specimen 1		Specimen 2		Specimen 3		Specimen 4		Specimen 5	
Troponin I	Elevated	Unelevated	Elevated	Unelevated	Elevated	Unelevated	Elevated	Unelevated	Elevated	Unelevated
Abbott i-STAT	1			1		1	1			1
J&J Vitros ECi luminometric	1		1		1		1			1
LifeSign MI	1			1	1		1			1
Nanogen Cardiac STATus	4		1	3	4		4			4
Tosoh AIA Pack fluorometric	1		1		1		1			1
Total Population	8	0	3	5	7	1	8	0	0	8
Flagging		*****				*****		*****	*****	

Due to a lack of participant consensus, Specimen 2 was not graded this event.

	Specimen 1		Specimen 2		Specimen 3		Specimen 4		Specimen 5	
Troponin T	Elevated	Unelevated	Elevated	Unelevated	Elevated	Unelevated	Elevated	Unelevated	Elevated	Unelevated
Roche CARDIAC T	3			3	1	2	3			3
Roche Trop T Sensitive	2			2		2	2			2
Total Population	5	0	0	5	1	4	5	0	0	5
Flagging		*****	*****		*****			*****	*****	

Correct responses are defined as those reflecting agreement among 80% or more of all participants.
Unacceptable responses are indicated by "****" on the Flagging line of each specimen.