AMERICAN ASSOCIATION OF BIOANALYSTS
205 West Levee St. Brownsville, TX 78520
(281) 436-5357 - (800) 234-5315 - Fax (713) 781-5008

PARTICIPANT STATISTICS **FECAL OCCULT BLOOD** **FIRST QUADRIMESTER 2010**

	Specimen 1		Specimen 2	
Name	Positive	Negative	Positive	Negative
Beck/Coult Hemmocult	91	20	2	110
Beck/Coult Hemmocult ICT	2	1		3
Beck/Coult Hemmocult SENSA	110	7	4	113
Cenogenics	2			2
Fischer Sure-Vue FOB	20	1		21
Fischer Sure-Vue ES	15		1	14
Helena ColoScreen	30	2	2	30
HemaPrompt	4	2	1	5
Hemosure Onestep	2	2		4
Henry Schein One-Step	1	3	2	2
Immunos Hema Screen Specifc		1		1
Immunostics Hema Screen	10	1	1	10
Inverness Clearview Ultra	21	11	1	31
Lab. Diag. QUIK-CULT	1			1
PolymedcoOC-Auto Micro 80		1		1
Propper Seracult	31	3		34
Propper Seracult Plus	12	1	2	11
PSS Select	9	1		10
PSS Select-ES	4		1	3
Quidel IFOB	11			11
Wampole FOB		1		1
Polymedco OC-Light		1		1
TOTAL POPULATION	376	59	17	419
Flagging		*****	*****	

Correct responses are defined as those reflecting agreement among 80% or more of all participants.
Unacceptable responses are indicated by "****" on the Flagging line of each specimen.

Due to a lack of participant consensus, Beckman Coulter Hemmocult and Inverness Clearview Ultra Specimen 1 were not evaluated.