AMERICAN ASSOCIATION OF BIOANALYSTS
205 West Levee St. Brownsville, TX 78520
(281) 436-5357 - (800) 234-5315 - Fax (713) 781-5008

PARTICIPANT STATISTICS

PREGNANCY

FIRST QUADRIMESTER 2010

Urine (hCG) Pregnancy	Specimen 1			Specimen 2			Specimen 3			Specimen 4			Specimen 5		
Method	Pos	Ind	Neg	Pos	Ind	Neg	Pos	Ind	Neg	Pos	Ind	Neg	Pos	Ind	Neg
Alpha Sci Instant-View Cmbc	6					6	6					6	6		
Beckman ICON 20/25 s/u	53				1	52	53			1	1	51	52		1
Beckman ICON II s/u	6					6	6					6	6		
CardinalHealth hCG Combc	32					32	32					32	32		
Diag Sp Pregna-Cert	20					20	20					20	20		
Fisher Sure-Vue hCG-STAT	31					31	31					31	31		
Fisher Sure-Vue s/u hCG	28		1	1		28	29					29	29		
Fisher Sure-Vue urine hCG	17					17	16		1	1		16	17		
Genzyme OSOM hCG-Combo	23					23	23					23	23		
Genzyme OSOM hCG-Urine	17					17	17					17	17		
Henry Schein OneStep	14					14	14					14	14		
Instant Tech One-Step	6			1		5	6					6	6		
Inverness Acceava Combc	17					17	17			1		16	16		1
McKesson Medi-Lab	22					22	22					22	22		
OneStep 3rd party, OEM	6					6	6					6	6		
OneStep 3rd party, OEM waiv	6					6	6					6	6		
PerMaxim OneStep	6					6	6					6	6		
Polymedco Poly stat hCG	12					12	12					12	12		
PSS Consult	6					6	6					6	6		
PSS Select Combo	13					13	13					13	13		
Qui QuickVue Urine/One-Step	93				1	92	93					92	92		
Quidel Quickvue OS Combc	30					30	30					30	30		
Quidel Quickvue+ OS Combc	48					48	48					48	48		
Stanbio QuPID PLUS	7					7	7					7	7		
Stanbio True 20 Plus	12					12	12					12	12		
Wampole Clearview Combo	9					9	9					9	9		
Wampole UCG (slide)	7		2		1	8	7		2		1	8	7		2
TOTAL POPULATION	610	0	3	2	3	607	609	0	3	3	2	606	607	0	4
Flagging		*****	*****	*****	*****			*****	*****	*****	*****			*****	*****

AMERICAN ASSOCIATION OF BIOANALYSTS
205 West Levee St. Brownsville, TX 78520
(281) 436-5357 - (800) 234-5315 - Fax (713) 781-5008

Serum (hCG) Pregnancy	Specimen 1			Specimen 2			Specimen 3			Specimen 4			Specimen 5		
Method	Pos	Ind	Neg	Pos	Ind	Neg	Pos	Ind	Neg	Pos	Ind	Neg	Pos	Ind	Neg
Beckman ICON 20/25 s/u	25					25	25					25	25		
CardinalHealth hCG Combo	22					22	22					22	22		
Fisher Sure-Vue hCG-STAT	14					14	14					14	14		
Fisher Sure-Vue s/u hCG	23					23	23					23	23		
Genzyme OSOM hCG-Combo	22					22	22					22	22		
Henry Schein OneStep	8					8	8					8	8		
Inverness Acceava Combo	26					26	26			1		25	25		1
McKesson Medi-Lab	8					8	8					8	8		
Polymedco Poly stat hCG	8					8	8					8	8		
PSS Select Combo	8					8	8					8	8		
Quidel QuickVue	9					9	9					9	9		
Quidel Quickvue OS Combo	20					20	20					20	20		
Quidel Quickvue+ OS Combo	35					35	35					35	35		
Wampole Clearview Combo	6					6	6					6	6		
TOTAL POPULATION	287	3	3	3	3	287	287	0	0	1	0	286	285	0	1
Flagging		*****	*****	*****	*****			*****	*****	*****	*****			*****	*****

hcg Concentration	2662 mIU/mL	<2 mIU/mL	2662 mIU/mL	<2 mIU/mL	2662 mIU/mL

Correct responses are defined as those reflecting agreement among 80% or more of all participants.
Unacceptable responses are indicated by "****" on the Flagging line of each specimen.