AMERICAN ASSOCIATION OF BIOANALYSTS
205 West Levee St. Brownsville, TX 78520
(281) 436-5357 - (800) 234-5315 - Fax (713) 781-5008

PARTICIPANT STATISTICS

PREGNANCY(qualitative hCG)

FIRST QUADRIMESTER 2010

	Specimen 1			Specimen 2			Specimen 3			Specimen 4			Specimen 5		
Name	Pos	Ind	Neg	Pos	Ind	Neg	Pos	Ind	Neg	Pos	Ind	Neg	Pos	Ind	Neg
Abbott MEIA / CMIA	2			2			2			2					2
ABI SureStep Combo	6			6			6			6					6
Acceva hCG Combo	7		1	8			7	1		7		1	1		7
Alfa Sci Instant-View Combo	5			5			5			5					5
American Diag Pro Combo	6			6			6			6					6
Beckman ICON 20, 25	21			21			21			21					21
Beckman Access	2			2			2			2					2
Beckman ICON II s/u	9			9			9			9					9
Cardinal Health/SP hCG Comb	17			17			17			17					17
Fisher Sure-Vue hCG-STAT	31	1	1	31	2		32	1		32	1		1		32
Fisher Sure-Vue s/u hCG	34			34			34			34					34
Genzyme Contrast	5			5			5			5					5
Genzyme OSOM hCG-Combo	25			25			25			25			1		24
Genzyme Rapid	3			3			3			3					3
Immunostics Detector Conbi	3			3			3			3					3
Inverness Clearview	5	1		5	1		5	1		5	1				6
Inverness Signify	4			4			4			4					4
LifeSign Status s/u hCG	2	1		2	1		2	1		2	1				3
McKesson hCG Combo	10			10			10			10					10
Mainline Confirms	2			2			2			2					2
OneStep 3rd party, OEM	7			7			7			7					7
Polymedco Poly stat hCG	9			9			9			9					9
Quidel QuickVue OS Combo	66			65		1	66			66					66
Quidel QuickVue+ OS Combo	25			25			25			25					25
SA Scientific SAS	3			3			3			3					3
PSS Select	2	1		2	1		2		1	2	1			1	2
Siemens Immulite	3			3			3			3					3
Stanbio One-Step	3			3			3			3					3
Stanbio QuPID PLUS	8			8			8			8				1	7
Syntron Quikpac II	2			2			2			2					2
Teco Diag visual	3			3			3			3					3
Wampole Clearview DUO	9			9			9			9					9
Wampole PreVue s/u	2			2			2			2					2
OneStep Com 3rd part non wv	2			2			2			2					2
TOTAL POPULATION	343	4	2	356	5	1	357	4	1	357	4	1	3	2	357
Flagging		*****	*****			*****		*****	*****		*****	*****	*****	*****	
hcg Concentration	3085 mIU/mL			235 mIU/mL			844 mIU/mL			1832 mIU/mL			<2 mIU/mL		

Correct responses are defined as those reflecting agreement among 80% or more of all participants.
Unacceptable responses are indicated by "****" on the Flagging line of each specimen.