URINE DRUG SCREENING
FIRST QUADRIMESTER 2010

AMERICAN ASSOCIATION OF BIOANALYSTS
205 West Levee St. Brownsville, TX 78520
(281) 436-5357 - (800) 234-5315 - Fax (713) 781-5008

ALCOHOL	Specimen 1 Present	Specimen 1 Absent	Specimen 2 Present	Specimen 2 Absent
Abbott AxSYM	6		6	
Beckman Synchron Emit	3		3	
iCup	3		3	
Microgenics DRI	9		9	
Q.E.D. Alcohol	3		3	
Roche UV alcohol	3		3	
Siemens Dimension	3		3	
Siemens V-Twin	3		3	
Syva Emit II, Emit II Plus	15		15	
Varian On-Site	3		3	
TOTAL POPULATION	68	1	67	2
Flagging		*****		*****
Assay Value (mg/dL)	*160*		*144*	

AMPHETAMINES	Specimen 1 Present	Specimen 1 Absent	Specimen 2 Present	Specimen 2 Absent
All chromatographic methods		4	4	
Abbott AxSYM		4	4	
ABM Rapid Drug Screen		3	3	
ACON Labs One Step		3	3	
ADP LiveSure		4	4	
Alfa Scientific InstantView		4	4	
ATN FasTox		9	9	
Beckman Coulter DXC		14	14	
Beckman Synchron Emit		3	3	
BioRad TOX/See		28	28	
Biosite Express		1	1	
Biosite TRIAGE		97	97	
BioView		4	4	
BMC QuickTox	1	3	3	1
Diag Reag Inc (DRI)		3	3	
iCassette		6	6	
iCup		21	20	1
Innovacon multi-CLIN	1	9	9	1
iScreen		4	4	
J&J Vitros		4	4	
Jant Pharmacal accutest		4	4	
MedTox - 5 Panel		4	4	
MedTox Profile series	1	70	70	1
MedTox Verdict-II	1	12	12	1
Microgenics DRI		12	12	
Noble Multi-CLIN		6	6	
OneStep 3rd party, OEM		5	5	
Redwood BioTech One Step		14	14	
Roche Abuscreen OnLine		3	3	
Roche Cobas		3	3	
Siemens Dimension		11	11	
Siemens Viva ser		4	4	
Siemens V-Twin		8	8	
Syva Emit II, Emit II Plus		24	24	
TOTAL POPULATION	4	443	442	5
Flagging	*****			*****
Assay Value (ng/mL)			*5003*	

BARBITURATES	Specimen 1 Present	Specimen 1 Absent	Specimen 2 Present	Specimen 2 Absent
All chromatographic methods		3		3
Abbott AxSYM		5		5
ABM Rapid Drug Screen		3		3
ACON Labs One Step		3		3
ADP LiveSure		3		3
Alfa Scientific InstantView		4		4
ATN FasTox		9		9
Beckman Coulter DXC		11		11
Beckman Synchron Emit		3		3
BioRad TOX/See		28		28
Biosite TRIAGE		95	1	95
BioView		4		4
BMC QuickTox	1	3	1	3
Diag Reag Inc (DRI)		3		3
iCassette		5		5
iCup		13		13
Innovacon multi-CLIN		9		9
J&J Vitros		4		4
Jant Pharmacal accutest		3		3
MedTox Profile series	1	68	1	68
MedTox Verdict-II	1	10	1	10
Microgenics DRI		11		11
Noble Multi-CLIN		3		3
Redwood BioTech One Step		12		12
Roche Abuscreen OnLine		3		3
Roche Cobas		3		3
Siemens Dimension		10		10
Siemens V-Twin		7		7
Syva Emit II, Emit II Plus		19		19
TOTAL POPULATION	3	390	4	390
Flagging	*****		*****	

BENZODIAZEPINES	Specimen 1 Present	Specimen 1 Absent	Specimen 2 Present	Specimen 2 Absent
All chromatographic methods	4			4
Abbott AxSYM	6			6
ABM Rapid Drug Screen	3			3
ACON Labs One Step	3			3
ADP LiveSure	3			3
Alfa Scientific InstantView	2	2		4
ATN FasTox	8	1		9
Beckman Coulter DXC	11			11
BioRad TOX/See	28			28
Biosite TRIAGE	95	1	1	95
BioView	4			4
BMC QuickTox	2	2	1	3
Diag Reag Inc (DRI)	3			3
iCassette	5			5
iCup	19			19
Innovacon multi-CLIN	9			9
iScreen	4			4
J&J Vitros	4			4
Jant Pharmacal accutest	3			3
MedTox Profile series	66	3	1	68
MedTox Verdict-II	11	1	1	11
Microgenics DRI	11			11
Noble Multi-CLIN	3			3
OneStep 3rd party, OEM	3			3
Redwood BioTech One Step	13			13
Roche Abuscreen OnLine	3			3
Roche Cobas	3			3
Siemens Dimension	11			11
Siemens V-Twin	6			6
Syva Emit II, Emit II Plus	21			21
TOTAL POPULATION	397	14	5	406
Flagging		*****	*****	
Assay Value (ng/mL)	*1650*			

CANNABINOIDS	Specimen 1 Present	Specimen 1 Absent	Specimen 2 Present	Specimen 2 Absent
All chromatographic methods		4		4
Abbott AxSYM		6		6
ABM Rapid Drug Screen		4		4
ACON Labs One Step		3		3
ADP LiveSure		4		4
Alfa Scientific InstantView		4		4
ATN FasTox		8		8
Beckman Coulter DXC		14		14
Beckman Synchron Emit		3		3
BioRad TOX/See		28		28
Biosite TRIAGE	2	94		96
BioView		3		3
BMC QuickTox	1	3	1	3
Diag Reag Inc (DRI)		3		3
iCassette		6		6
iCup	1	19		20
Innovacon multi-CLIN	1	9	1	9
Instant Tech. One Step		4		4
iScreen		5		5
J&J Vitros		4		4
Jant Pharmacal accutest		4		4
MedTox - 5 Panel		5		5
MedTox Profile series	1	70	1	70
MedTox Verdict-II	1	11	1	11
Microgenics DRI		13		13
Noble Multi-CLIN		6		6
OneStep 3rd party, OEM		5		5

AMERICAN ASSOCIATION OF BIOANALYSTS
205 West Levee St. Brownsville, TX 78520
(281) 436-5357 - (800) 234-5315 - Fax (713) 781-5008

	Specimen 1		Specimen 2	
	Present	Absent	Present	Absent
Redwood BioTech One Step		19		19
Roche Abuscreen OnLine		3		3
Roche Cobas		3		3
Siemens Dimension		11		11
Siemens Viva ser		8		8
Siemens V-Twin		9		9
Syva Emit DAU		6		6
Syva Emit II, Emit II Plus		56		56
TOTAL POPULATION	7	490	4	493
Flagging	*****		*****	
Assay Value (ng/mL)				

COCAINE METABOLITE	Specimen 1 Present	Specimen 1 Absent	Specimen 2 Present	Specimen 2 Absent
All chromatographic methods		4		4
Abbott AxSYM		6		6
ABM Rapid Drug Screen		4		4
ACON Labs One Step		3		3
ADP LiveSure		4		4
Alfa Scientific InstantView		5		5
ATN FasTox		9		9
Beckman Coulter DXC		14		14
Beckman Synchron Emit		3		3
BioRad TOX/See		28		28
Biosite TRIAGE		97		97
BioView		4		4
BMC QuickTox	1	3	2	2
Diag Reag Inc (DRI)		3		3
iCassette		6		6
iCup		22		22
Innovacon multi-CLIN	1	9	1	9
Instant Tech. One Step		4	1	3
iScreen		5		5
J&J Vitros		4		4
Jant Pharmacal accutest		4		4
MedTox - 5 Panel		5		5
MedTox Profile series	1	70	1	70
MedTox Verdict-II	1	11	1	11
Microgenics DRI		13		13
Noble Multi-CLIN		6		6
OneStep 3rd party, OEM		5		5
Redwood BioTech One Step		18		18
Roche Abuscreen OnLine		3		3
Roche Cobas		3		3
Roche/VarianOnTrak TesTstik		3		3
Siemens Dimension		11		11
Siemens Viva ser		8		8
Siemens V-Twin		10		10
Syva Emit DAU		6		6
Syva Emit II, Emit II Plus		54		54
TOTAL POPULATION	4	500	6	48
Flagging	*****		*****	
Assay Value (ng/mL)				

MDMA (ECSTACY)	Specimen 1 Present	Specimen 1 Absent	Specimen 2 Present	Specimen 2 Absent
ATN FasTox		9	9	
BioRad TOX/See		18	18	
iCup		8	8	
Innovacon multi-CLIN		5	5	
MedTox Profile series	1	8	6	3
Microgenics DRI		4	4	
Noble Multi-CLIN		3	3	
Redwood BioTech One Step		6	6	
Siemens V-Twin		3	3	
TOTAL POPULATION	2	90	87	5
Flagging	*****			*****
Assay Value (ng/mL)			*4470*	

METHADONE	Specimen 1 Present	Specimen 1 Absent	Specimen 2 Present	Specimen 2 Absent
All chromatographic methods		3		3
Abbott AxSYM		5		5
ADP LiveSure		3		3
ATN FasTox		9		9
Beckman Coulter DXC		7		7
BioRad TOX/See		21		21
Biosite TRIAGE		66		65
BioView		3		3
BMC QuickTox	1	3	1	3
iCassette		3		3
iCup		8		8
Innovacon multi-CLIN		6		6
J&J Vitros		3		3
MedTox Profile series	2	53	1	54
MedTox Verdict-II	1	3	1	3
Microgenics DRI		9		9
Pharmatech QuickScreen		2	1	1
Redwood BioTech One Step		13		13
Siemens Dimension		4		4
Siemens V-Twin		4		4
Syva Emit II, Emit II Plus		13		13
TOTAL POPULATION	4	280	5	278
Flagging	*****		*****	
Assay Value (ng/mL)				

METHADONE METABOLITE	Specimen 1 Present	Specimen 1 Absent	Specimen 2 Present	Specimen 2 Absent
iCup	3			3
MedTox Profile series		4	1	3
Microgenics DRI	6	1		7
TOTAL POPULATION	12	9	1	20
Flagging			*****	
Assay Value (ng/mL)	*920*			

Due to a lack of participant consenesus Methadone Metabolite Specimen 1 will not be graded this event.

METHAMPHETAMINE	Specimen 1 Present	Specimen 1 Absent	Specimen 2 Present	Specimen 2 Absent
All chromatographic methods		4	4	
ABM Rapid Drug Screen		3	3	
ADP LiveSure		3	3	
Alfa Scientific InstantView		5	5	
ATN FasTox		9	9	
BioRad TOX/See		28	28	
Biosite TRIAGE	1	67	67	1
BioView		4	4	
BMC QuickTox	1	3	3	1
iCassette		5	5	
iCup		21	21	
Innovacon multi-CLIN	1	5	5	1
Instant Tech. One Step		3	2	1
iScreen		5	5	
Jant Pharmacal accutest		3	3	
MedTox Profile series	1	57	58	
MedTox Verdict-II	1	4	4	1
Noble Multi-CLIN		4	4	
OneStep 3rd party, OEM		4	4	
Redwood BioTech One Step		13	13	
TOTAL POPULATION	5	273	270	8
Flagging	*****			*****
Assay Value (ng/mL)			*2370*	

METHAQUALONE	Specimen 1 Present	Specimen 1 Absent	Specimen 2 Present	Specimen 2 Absent
Beckman Coulter DXC	4			4
iCup	3			3
Integrated E-Z Split		1		1
Microgenics DRI	6			6
Siemens V-Twin	3			3
Syva Emit II, Emit II Plus	3			3
TOTAL POPULATION	26	2	0	26
Flagging		*****	*****	
Assay Value (ng/mL)	*1800*			

OPIATES	Specimen 1 Present	Specimen 1 Absent	Specimen 2 Present	Specimen 2 Absent
All chromatographic methods		4	4	
Abbott AxSYM		6	6	
ABM Rapid Drug Screen		4	4	

AMERICAN ASSOCIATION OF BIOANALYSTS
205 West Levee St. Brownsville, TX 78520
(281) 436-5357 - (800) 234-5315 - Fax (713) 781-5008

	Specimen 1		Specimen 2	
	Present	Absent	Present	Absent
ACON Labs One Step		3	3	
ADP LiveSure		4	4	
Alfa Scientific InstantView		5	5	
ATN FasTox		9	9	
Beckman Coulter DXC		14	14	
Beckman Synchron Emit		3	3	
BioRad TOX/See		28	27	1
Biosite TRIAGE	1	95	95	1
BioView		4	4	
BMC QuickTox	1	3	3	1
Diag Reag Inc (DRI)		3	3	
iCassette		6	6	
Innovacon multi-CLIN	1	9	8	2
Instant Tech. One Step		4	4	
iCup		21	19	2
iScreen		5	5	
Jant Pharmacal accutest		4	4	
J&J Vitros		4	4	
MedTox - 5 Panel		4	4	
MedTox Profile series	1	70	70	1
MedTox Verdict-II	1	12	12	1
Microgenics DRI		13	13	
Noble Multi-CLIN		6	6	
OneStep 3rd party, OEM		5	5	
Redwood BioTech One Step		18	18	
Roche Abuscreen OnLine		3	3	
Roche Cobas		3	3	
Roche/VarianOnTrak TesTstik		3	3	
Siemens Dimension		11	11	
Siemens Viva ser		8	8	
Siemens V-Twin		10	10	
Syva Emit DAU		6	6	
Syva Emit II, Emit II Plus		55	55	
TOTAL POPULATION	5	499	495	9
Flagging	*****			*****
Assay Value (ng/mL)			*3490*	

	Specimen 1		Specimen 2	
OXYCODONE	Present	Absent	Present	Absent
All chromatographic methods		3		3
ATN FasTox		9	5	4
BioRad TOX/See		7		7
BMC QuickTox	1	2	2	1
Innovacon multi-CLIN		7		7
iCup	1	9		10
MedTox Profile series		38	4	34
MedTox Verdict-II		4	2	2
Microgenics DRI		4		4
Noble Multi-CLIN		3		3
Redwood BioTech One Step		7		7
TOTAL POPULATION	3	110	0 14	99
Flagging	*****		*****	

	Specimen 1		Specimen 2	
PHENCYCLIDINE	Present	Absent	Present	Absent
All chromatographic methods	3			3
Abbott AxSYM	4			4
ADP LiveSure	3			3
Alfa Scientific InstantView	2	1		3
ATN FasTox	5			5
Beckman Coulter DXC	5			5
BioRad TOX/See	10			10
Biosite TRIAGE	58	1	2	57
BioView	3			3
BMC QuickTox	3	1	1	3
iCassette	5			5
iCup	8	1	1	8
Innovacon multi-CLIN	6	1	1	6
MedTox Profile series	36	2		38
MedTox Verdict-II	6	1	1	6
Microgenics DRI	7			7
Noble Multi-CLIN	4			4
Redwood BioTech One Step	12			12
Roche Cobas	3			3
Siemens Dimension	8			8
Siemens V-Twin	7			7

	Specimen 1		Specimen 2	
	Present	Absent	Present	Absent
Syva Emit II, Emit II Plus	16			16
TOTAL POPULATION	248	11	6	253
Flagging		*****	*****	
Assay Value (ng/mL)	*190*			

	Specimen 1		Specimen 2	
PROPOXYPHENE	Present	Absent	Present	Absent
Beckman Coulter DXC		5	5	
Biosite TRIAGE		3	3	
Innovacon multi-CLIN		6	6	
iCup	1	8	8	1
MedTox Profile series		50	50	
MedTox Verdict-II	1	4	4	1
Microgenics DRI		11	11	
Redwood BioTech One Step		5	5	
Siemens V-Twin		4	4	
Syva Emit II, Emit II Plus	1	9	10	
TOTAL POPULATION	3	129	129	3
Flagging	*****			*****
Assay Value (ng/mL)			*1900*	

	Specimen 1		Specimen 2	
TRICYCLIC ANTIDEPRES	Present	Absent	Present	Absent
ATN FasTox	4	1		5
BioRad TOX/See	17			17
Biosite TRIAGE	53	1	1	53
BioView	3			3
Innovacon multi-CLIN	5			5
iCup	5			5
MedTox Profile series	59	1		60
MedTox Verdict-II	3	1	1	3
Redwood BioTech One Step	4			4
TOTAL POPULATION	153	4	2	155
Flagging		*****	*****	
Assay Value (ng/mL)	*3950*			