



AMERICAN ASSOCIATION OF BIOANALYSTS
 205 West Levee St. Brownsville, TX 78520
 (281) 436-5357 - (800) 234-5315 - Fax (713) 781-5008

PARTICIPANT STATISTICS

COMPREHENSIVE CHEMISTRY

Chemistry Q1 2010

Name	Line No.	Specimen 1			Specimen 2			Specimen 3			Specimen 4			Specimen 5			No. Labs					
		Range & Type	Mean	SD	Range & Type	Mean	SD	Range & Type	Mean	SD	Range & Type	Mean	SD	Range & Type	Mean	SD						
T-Uptake Ratio to Normal																						
Initial Grouping by Reagent and Instrument																						
Abbott FPIA & Abbott AxSYM	1	0.33-0.53	S	0.432	0.03	0.24-0.45	S	0.346	0.04	0.54-0.73	S	0.637	0.03	0.34-0.55	S	0.444	0.03	0.15-0.33	S	0.236	0.03	14
Abbott FPIA & Abbott IMx	2	0.32-0.55	S	0.436	0.04	0.25-0.45	S	0.349	0.03	0.54-0.75	S	0.646	0.03	0.34-0.54	S	0.440	0.03	0.16-0.33	S	0.245	0.03	14
Initial Grouping by Reagent																						
Abbott FPIA	3	0.33-0.54	S	0.434	0.04	0.25-0.45	S	0.347	0.03	0.54-0.74	S	0.642	0.03	0.34-0.54	S	0.442	0.03	0.15-0.33	S	0.240	0.03	28
Initial Grouping by Sensitivity or Principle																						
Fluorescence polariz (FPIA)	4	0.33-0.54	S	0.434	0.04	0.25-0.45	S	0.347	0.03	0.54-0.74	S	0.642	0.03	0.34-0.54	S	0.442	0.03	0.15-0.33	S	0.240	0.03	28
Other	5	0.12-0.99	S	0.559	0.14	0.09-0.95	S	0.522	0.14	0.00-0.74	S	0.318	0.14	0.20-0.76	S	0.482	0.09	0.00-1.29	S	0.620	0.22	12
Total Population																						
Whole Population	6	0.21-0.71	S	0.464	0.08	0.15-0.62	S	0.383	0.08	0.47-0.78	S	0.625	0.05	0.29-0.62	S	0.453	0.06	0.18-0.34	S	0.260	0.03	43