



**PROFICIENCY TESTING SERVICE
AMERICAN ASSOCIATION OF BIOANALYSTS**

205 West Levee St. ♦ Brownsville, TX 78520-5596
800-234-5315 ♦ 281-436-5357 ♦ Fax 713-781-5008

FIRST QUADRIMESTER 2011

Alcohol - Acetone

Acetone	Specimen 1				Specimen 2				Specimen 3				Specimen 4				Specimen 5				
	Neg	Small (20 mg/dL)	Moderate (30-40 mg/dL)	Large (80-100 mg/dL)	Neg	Small (20 mg/dL)	Moderate (30-40 mg/dL)	Large (80-100 mg/dL)	Neg	Small (20 mg/dL)	Moderate (30-40 mg/dL)	Large (80-100 mg/dL)	Neg	Small (20 mg/dL)	Moderate (30-40 mg/dL)	Large (80-100 mg/dL)	Neg	Small (20 mg/dL)	Moderate (30-40 mg/dL)	Large (80-100 mg/dL)	
Bayer Acetest	7				7					7				7						6	1
Method not reported	71				71					63	8		71						11	50	9
Total Population	78	0	0	0	78	0	0	0	0	70	8	0	78	0	0	0	1	11	56	10	
<i>Flagging</i>		***	***	***		***	***	***	***					***	***	***	***				

Correct responses are defined as those reflecting agreement among 80% or more of all participants.
Unacceptable responses are indicated by "*****" on the Flagging line of each specimen.