



**PROFICIENCY TESTING SERVICE
AMERICAN ASSOCIATION OF BIOANALYSTS**

205 West Levee St. ♦ Brownsville, TX 78520-5596
800-234-5315 ♦ 281-436-5357 ♦ Fax 713-781-5008

FIRST QUADRIMESTER 2011

Serum Pregnancy - Qualitative HCG

Name	Specimen 1			Specimen 2			Specimen 3			Specimen 4			Specimen 5		
	Pos	Ind	Neg	Pos	Ind	Neg	Pos	Ind	Neg	Pos	Ind	Neg	Pos	Ind	Neg
ABI SureStep Combo	5			5			5					5	5		
Acceva hCG Combo	7			7			7					7	7		
Alfa Sci Instant-View Combo	8			8			8					8	8		
American Diag Pro Combo	6			6			6					6	6		
Beckman Access	2			2			2					2	2		
Beckman ICON 20, 25	14			14			14					14	14		
Beckman ICON II s/u	8			8			8					8	8		
Bio-Med Double-Chek s/u	2			2			2					2	2		
Cardinal Health/SP hCG Comb	19			19			19					19	19		
Fisher Sure-Vue hCG-STAT	28			28			28					28	28		
Fisher Sure-Vue s/u hCG	25	1		25	1		25	1			1	25	24	1	
Genzyme Contrast	3			3			3					3	3		
Genzyme OSOM hCG-Combo	21	1		21	1		21	1				22	21	1	
Immunostics Detector Conbi	3			3			3					3	3		
Inverness Clearview	8			8			8					8	8		
Inverness Signify	2			2			2					2	2		
LifeSign Status s/u hCG	3			3			3					3	3		
Mainline Confirms	2			2			2					2	2		
McKesson hCG Combo	6			6			6					6	6		
Meridian ImmunoCard STAT	4			4			4					4	4		
OneStep 3rd party, OEM	5			5			5					5	5		
OneStep Com 3rd part non wv	5			5			5					5	5		
Polymedco Poly stat hCG	6			6			6					6	6		
Quidel QuickVue OS Combo	60	1		60	1		60	1				61	59	1	1
Quidel QuickVue+ OS Combo	24			24			24					24	24		
SA Scientific SAS	5			5			5					5	5		
Select Diag hCG Combo	3			3			3					3	3		
Siemens Immulite	3			3			3					3	3		
Stanbio One-Step	3			3			3					3	3		
Stanbio QuPID PLUS	6			6			6					6	6		
Teco Diag visual	2			2			2					2	2		
Wampole Clearview DUO	4	1		4	1		4	1				5	4	1	
Wampole PreVue s/u	2			2			2					2	2		
Total Population	352	5	0	353	5		353	5			1	356	350	5	1
Flagging		***	***			***		***	***	***	***	***		***	***

hcg Concentration

1862 mIU/mL

792 mIU/mL

2228 mIU/mL

<3 mIU/mL

2858 mIU/mL

Correct responses are defined as those reflecting agreement among 80% or more of all participants. Unacceptable responses are indicated by "*****" on the Flagging line of each specimen.

Basic Chemistry