



D-Dimer

Name	Line No.	Specimen 1			Specimen 2			No. of Labs		
		Range & Type	Mean	SD	Range & Type	Mean	SD			
<b>D-Dimer, quantitative</b>										
<b>Initial Grouping by Method</b>										
Biosite Triage	1	480.0 - 1302.0	S	891.2	137.0	468.0 - 935.0	S	701.5	77.9	109
IL ACL Elite	2	531.0 - 793.0	S	662.1	43.6	420.0 - 673.0	S	546.2	42.2	10
Siemens Stratus CS	3	1737.0 - 2699.0	S	2217.8	160.4	1282.0 - 1941.0	S	1611.5	110.0	26
Siemens Sysmex Advanced	4	1575.0 - 3385.0	S	2480.0	301.8	1415.0 - 2835.0	S	2125.0	236.7	17
<b>Initial Grouping by Sensitivity or Principle</b>										
low recovery	5	480.0 - 1302.0	S	891.2	137.0	468.0 - 935.0	S	701.5	77.9	109
mid recovery	6	1575.0 - 3385.0	S	2480.0	301.8	1415.0 - 2835.0	S	2125.0	236.7	17
high recovery	7	0.0 - 4377.0	S	1994.2	794.3	0.0 - 3713.0	S	1516.5	732.2	41
<b>Total Population</b>										
Whole Population	8	0.0 - 3502.0	S	1243.2	753.1	0.0 - 2823.0	S	977.5	615.1	204