



High Sensitivity C-Reactive Protein

Name	Line No.	Specimen 1			Specimen 2			No. of Labs		
		Range & Type	Mean	SD	Range & Type	Mean	SD			
High Sensitivity C-Reactive Protein										
Initial Grouping by Method										
Beckman Unicel	1	0.0 - 14.29	S	6.360	2.643	0.0 - 9.99	S	4.441	1.851	14
Carolina immunoturbid	2	6.62 - 9.21	S	7.919	0.432	4.0 - 7.17	S	5.586	0.528	13
Siemens CCRP	3	5.46 - 11.62	S	8.542	1.026	4.56 - 8.06	S	6.311	0.584	15
Siemens Immulite 2000	4	6.45 - 8.45	S	7.451	0.334	4.42 - 5.71	S	5.067	0.215	10
Initial Grouping by Sensitivity or Principle										
All wide range methods	5	1.41 - 13.19	S	7.297	1.963	0.94 - 9.11	S	5.028	1.362	16
All Roche HS methods	6	5.39 - 7.85	S	6.622	0.409	4.06 - 5.42	S	4.742	0.226	21
All direct quantitative meth	7	1.92 - 12.39	S	7.157	1.746	1.57 - 8.6	S	5.085	1.171	107
All quantitations by dilutn	8	4.87 - 10.89	S	7.881	1.003	3.13 - 8.1	S	5.619	0.828	38
Total Population										
Whole Population	9	5.44 - 9.55	S	7.496	0.686	3.57 - 6.96	S	5.264	0.565	186