



**PROFICIENCY TESTING SERVICE
AMERICAN ASSOCIATION OF BIOANALYSTS**

205 West Levee St. • Brownsville, TX 78520-5596
800-234-5315 • 281-436-5357 • Fax 713-781-5008

FIRST QUADRIMESTER 2011

Urine Pregnancy

| Name | Specimen 1 | | | Specimen 2 | | | Specimen 3 | | | Specimen 4 | | | Specimen 5 | | |
|-----------------------------|------------|----------|------------|------------|----------|------------|------------|-----------|----------|------------|-----------|----------|------------|----------|------------|
| | Pos | Ind | Neg | Pos | Ind | Neg | Pos | Ind | Neg | Pos | Ind | Neg | Pos | Ind | Neg |
| Alpha Sci Instant-View Cmbo | | | 12 | | | 12 | 12 | | | 12 | | | | | 12 |
| Beckman ICON 20/25 s/u | | | 45 | | | 45 | 44 | 1 | | 43 | 1 | 1 | | | 45 |
| Beckman ICON II s/u | | | 8 | | | 8 | 7 | 1 | | 7 | | | | | 8 |
| Cardinal Health SP | | | 7 | | | 7 | 7 | | | 7 | | | | | 7 |
| CardinalHealth hCG Combo | | | 22 | | | 22 | 22 | | | 22 | | | | | 22 |
| Diag Sp Pregna-Cert | | | 18 | | | 18 | 18 | | | 18 | | | | | 18 |
| Diag Test Group Clarity | | | 3 | | | 3 | 3 | | | 3 | | | | | 3 |
| Fisher Sure-Vue hCG-STAT | | | 33 | | | 33 | 32 | 1 | | 32 | 1 | | 1 | | 32 |
| Fisher Sure-Vue s/u hCG | | | 21 | | | 21 | 20 | 1 | | 20 | 1 | | | | 21 |
| Fisher Sure-Vue urine hCG | | | 12 | | | 12 | 12 | | | 12 | | | 1 | | 11 |
| Genzyme OSOM hCG-Combo | | | 19 | | | 19 | 19 | | | 19 | | | | 1 | 18 |
| Genzyme OSOM hCG-Urine | | | 17 | | | 17 | 17 | | | 17 | | | | | 17 |
| Germaine AimStep Combo. ur | | | 3 | | | 3 | 3 | | | 3 | | | | | 3 |
| Henry Schein OneStep | | | 9 | | | 9 | 8 | 1 | | 8 | 1 | | | | 9 |
| Inverness Acceava Combo | | | 14 | | | 14 | 14 | | | 14 | | | | | 14 |
| Inverness BioStar Acceav II | | | 3 | | | 3 | 3 | | | 3 | | | | | 3 |
| Inverness Clearview Ultra | | | 4 | | | 4 | 4 | | | 4 | | | | | 4 |
| Lifesign Status s/u hCG | | | 5 | | | 5 | 5 | | | 5 | | | | 1 | 4 |
| Mainline Confirms hCG | | | 3 | | | 3 | 3 | | | 3 | | | | | 3 |
| McKesson Medi-Lab | | | 25 | | | 25 | 25 | | | 25 | | | | | 25 |
| OneStep 3rd party. OEM | | | 4 | | | 4 | 4 | | | 4 | | | | | 4 |
| OneStep 3rd party. OEM waiv | | | 6 | | | 6 | 6 | | | 6 | | | | | 6 |
| PerMaxim OneStep | | | 6 | | | 6 | 5 | 1 | | 5 | 1 | | | | 6 |
| Polymedco Poly stat hCG | | | 11 | | | 11 | 11 | | | 11 | | | | | 11 |
| PSS Consult | | | 12 | | | 12 | 12 | | | 12 | | | | | 12 |
| PSS Select Combo | | | 10 | | | 10 | 10 | | | 10 | | | | | 10 |
| Qui QuickVue Urine/One-Step | | | 66 | | | 66 | 65 | 1 | | 65 | 1 | | | | 66 |
| Quidel Quickvue OS Combo | | | 37 | | | 37 | 36 | 1 | | 36 | 1 | | | | 37 |
| Quidel Quickvue+ OS Combo | | | 32 | | | 32 | 29 | 3 | | 29 | 3 | | | | 32 |
| SA Scientific SAS | | | 3 | | | 3 | 3 | | | 3 | | | | | 3 |
| Siemens Clinitek Status | | | 4 | | | 4 | 4 | | | 4 | | | | | 4 |
| Stanbio QuPID | | | 4 | | | 4 | 4 | | | 4 | | | | | 4 |
| Stanbio QuPID PLUS | | | 7 | | | 7 | 7 | | | 7 | | | | | 7 |
| Stanbio True 20 Plus | | | 12 | | | 12 | 12 | | | 12 | | | | | 12 |
| Wampole Clearview Combo | | | 5 | | | 5 | 4 | 1 | | 4 | 1 | | | | 5 |
| Wampole Clearview II | | | 3 | | | 3 | 3 | | | 3 | | | | | 3 |
| Total Population | 1 | 0 | 552 | 1 | 0 | 552 | 539 | 13 | 0 | 538 | 12 | 1 | 2 | 2 | 548 |
| Flagging | *** | *** | | *** | *** | | *** | *** | *** | *** | *** | *** | *** | *** | *** |

Serum Pregnancy

| Name | Specimen 1 | | | Specimen 2 | | | Specimen 3 | | | Specimen 4 | | | Specimen 5 | | |
|-----------------------------|------------|----------|------------|------------|----------|------------|------------|----------|------------|------------|----------|----------|------------|----------|------------|
| | Pos | Ind | Neg | Pos | Ind | Neg | Pos | Ind | Neg | Pos | Ind | Neg | Pos | Ind | Neg |
| Alpha Sci Instant-View Cmbo | 1 | | 6 | 1 | | 6 | 7 | | | 6 | | 1 | 1 | | 6 |
| Beckman ICON 20/25 s/u | | | 26 | | | 26 | 25 | 1 | | 25 | 1 | | | | 26 |
| Beckman ICON II s/u | | | 5 | | | 5 | 5 | | | 5 | | | | | 5 |
| CardinalHealth hCG Combo | | | 17 | | | 17 | 16 | 1 | | 16 | 1 | | | | 17 |
| Fisher Sure-Vue hCG-STAT | | | 13 | | | 13 | 13 | | | 13 | | | | | 13 |
| Fisher Sure-Vue s/u hCG | | | 21 | | | 21 | 20 | 1 | | 20 | 1 | | | | 21 |
| Genzyme OSOM hCG-Combo | 1 | | 22 | 1 | | 22 | 22 | 1 | | 21 | 1 | 1 | 1 | | 22 |
| Henry Schein OneStep | | | 7 | | | 7 | 7 | | | 7 | | | | | 7 |
| Inverness Acceava Combo | | | 24 | | | 24 | 23 | 1 | | 23 | 1 | | | | 24 |
| McKesson Medi-Lab | | | 9 | | | 9 | 8 | 1 | | 8 | 1 | | | | 9 |
| OneStep 3rd party. OEM | | | 3 | | | 3 | 3 | | | 3 | | | | | 3 |
| Polymedco Poly stat hCG | | | 7 | | | 7 | 7 | | | 7 | | | | | 7 |
| PSS Consult | | | 3 | | | 3 | 3 | | | 3 | | | | | 3 |
| PSS Select Combo | | | 6 | | | 6 | 6 | | | 6 | | | | | 6 |
| Quidel QuickVue | | | 4 | | | 4 | 4 | | | 4 | | | | | 4 |
| Quidel Quickvue OS Combo | | | 21 | | | 21 | 21 | | | 21 | | | | | 21 |
| Quidel Quickvue+ OS Combo | | | 21 | | | 21 | 21 | 1 | | 21 | 1 | | | | 22 |
| SA Scientific SAS | | | 4 | | | 4 | 4 | | | 4 | | | | | 4 |
| Siemens Immulite | | 1 | 3 | | 1 | 3 | 4 | | | 4 | | | | 1 | 3 |
| Wampole Clearview Combo | | | 3 | | | 3 | 3 | | | 3 | | | | | 3 |
| Total Population | 2 | 1 | 251 | 2 | 1 | 251 | 248 | 7 | *** | 246 | 7 | 2 | 2 | 2 | 251 |
| Flagging | *** | *** | | *** | *** | | *** | *** | *** | *** | *** | *** | *** | *** | *** |

hcg Concentration

<1 mIU/mL

<1 mIU/mL

3030 mIU/mL

3030 mIU/mL

<1 mIU/mL

Correct responses are defined as those reflecting agreement among 80% or more of all participants.
Unacceptable responses are indicated by "*****" on the Flagging line of each specimen.