

Name	Line No.	Specimen 1			Specimen 2			No. of Labs
		Range & Type	Mean	SD	Range & Type	Mean	SD	
Dimension Xpand	3	178.0 - 240.0 P	209.1	12.0	180.0 - 244.0 P	211.8	8.1	19
Initial Grouping by Reagent								
Beckman Synchron	4	201.0 - 272.0 P	236.8	8.9	201.0 - 272.0 P	236.3	8.4	29
Carolina	5	199.0 - 270.0 P	234.7	8.8	200.0 - 271.0 P	235.7	8.9	10
J&J Vitros	6	187.0 - 254.0 P	220.6	7.2	187.0 - 253.0 P	219.8	11.9	14
Beckman Olympus	7	195.0 - 263.0 P	229.1	16.7	193.0 - 261.0 P	227.2	17.4	20
Siemens DCA 2000	8	200.0 - 270.0 P	234.8	7.9	198.0 - 268.0 P	232.6	7.9	49
Siemens Dimension	9	183.0 - 248.0 P	215.2	14.2	185.0 - 250.0 P	217.7	12.9	47
Siemens DCA Vantage	10	199.0 - 269.0 P	234.2	10.5	197.0 - 267.0 P	232.3	13.7	13
Initial Grouping by Sensitivity or Principle								
Jaffe/modified Jaffe	11	192.0 - 260.0 P	226.1	14.9	192.0 - 260.0 P	226.1	14.4	230
Total Population								
Whole Population	12	193.0 - 261.0 P	227.1	17.2	193.0 - 261.0 P	227.2	16.8	250