



**PROFICIENCY TESTING SERVICE
AMERICAN ASSOCIATION OF BIOANALYSTS**

205 West Levee St. ♦ Brownsville, TX 78520-5596
800-234-5315 ♦ 281-436-5357 ♦ Fax 713-781-5008

FIRST QUADRIMESTER 2016

Serum Pregnancy - Qualitative HCG

Name	Specimen 1			Specimen 2			Specimen 3			Specimen 4			Specimen 5			
	Neg	Pos	Ind	Neg	Pos	Ind	Neg	Pos	Ind	Neg	Pos	Ind	Neg	Pos	Ind	
Alere	7				7			7			7			7		
Alfa Scientific InstantView	12				12			12			12			12		
Beckman ICON 20, 25	8				8			8			8			8		
CardinalHealth hCG Combo	6				6			6			6			6		
Consult hCG Combo	10				10			10			10			10		
Fisher Sure-Vue hCG-STAT	7				7			7			7			7		
Fisher Sure-Vue s/u hCG	14				14			14			14			14		
OneStep 3rd party, OEM	5				4	1		5			5			5		
Polymedco Poly stat hCG	4				4			4			4			4		
Quidel QuickVue OS Combo	9	1		1	9			10			10			10		
Quidel QuickVue+ OS Combo	5	1			6			6			6			6		
Sekisui OSOM hCG-Combo	37				36	1		37			37		1	36		
Total Population	170	4	0	1	171	2	0	174	0	0	174	0	1	172	0	0
Flagging		***	***	***		***	***			***			***		***	***

hcg Concentration <0.7 mIU/mL 40 mIU/mL 2370 mIU/mL 1423 mIU/mL 660 mIU/mL

Correct responses are defined as those reflecting agreement among 80% or more of all participants.
Unacceptable responses are indicated by "****" on the Flagging line for each specimen.