



**PROFICIENCY TESTING SERVICE
AMERICAN ASSOCIATION OF BIOANALYSTS**

205 West Levee St. ♦ Brownsville, TX 78520-5596
800-234-5315 ♦ 281-436-5357 ♦ Fax 713-781-5008

Gastric Occult Blood

FIRST QUADRIMESTER 2016

Gastric Occult Blood

Name
Beckman Coulter Gastrocult
Total Population
Flagging

Specimen 1

Neg	Pos
	10
0	11

Specimen 2

Neg	Pos
9	1
10	1

pH - Gastric

Name
Visual Color Comparison/pH Paper
Total Population
Flagging

Specimen 1

2.0	3.0	4.0	5.0	>5.0
	4			
1	6	0	0	0
			***	***

Specimen 2

2.0	3.0	4.0	5.0	>5.0
	2	1		1
0	4	1	0	2

*Specimen 2 was not evaluated this event. The sample pH was 6.0

Correct responses are defined as those reflecting agreement among 80% or more of all participants. Unacceptable responses are indicated by "*****" on the Flagging line of each specimen.