

Siemens Dimension	4	10.1 - 15.1	P 12.58	0.34	0 - 0.5	C 0.11	0.03	-	-	-	10
Initial Grouping bySensitivityor Principle											
Diazonium ion	5	10.6 - 15.9	P 13.22	0.26	0 - 0.4	C 0.03	0.04	-	-	-	20
Diazo/caffeine-benzoate	6	10.1 - 15.1	P 12.58	0.32	0 - 0.5	C 0.12	0.04	-	-	-	12
Diazo/surfactant	7	10.1 - 15.1	P 12.58	0.23	0 - 0.6	C 0.2	0.0	-	-	-	4
Total Population											
Whole Population	8	10.1 - 15.1	P 12.57	1.56	0 - 0.5	C 0.07	0.07	-	-	-	39

Bilirubin, Total supplemental

Initial Grouping byReagent and Instrument											
Abbott Architect & Abbott Architect c, ci, i	1	9.2 - 13.8	P 11.51	0.52	0 - 0.5	C 0.11	0.03	-	-	-	8
Beckman Coulter & Beck Coult Unicel DXC	2	9.8 - 14.8	P 12.3	0.07	0 - 0.7	C 0.25	0.09	-	-	-	5
Beckman Olympus & Beck Olym AU 400/600/5400	3	10.4 - 15.6	P 13.0	0.51	0 - 0.6	C 0.2	0.0	-	-	-	4
Ortho Vitros & Ortho Vitros not DT or ECi	4	10.4 - 15.5	P 12.94	0.46	0 - 0.5	C 0.12	0.07	-	-	-	8
Siemens Dimension & Siemens Dimension EXL	5	9.9 - 14.9	P 12.39	0.3	0 - 0.5	C 0.1	0.0	-	-	-	13
Siemens Dimension & Siemens Dimension Xpand	6	9.9 - 14.8	P 12.35	0.44	0 - 0.5	C 0.1	0.0	-	-	-	7
Initial Grouping byReagent											
Abbott Architect	7	9.2 - 13.8	P 11.51	0.52	0 - 0.5	C 0.11	0.03	-	-	-	8
Beckman Coulter	8	9.8 - 14.8	P 12.3	0.07	0 - 0.7	C 0.25	0.09	-	-	-	5
Beckman Olympus	9	10.4 - 15.6	P 12.98	0.42	0 - 0.6	C 0.2	0.0	-	-	-	6
Ortho Vitros	10	10.2 - 15.3	P 12.75	0.45	0 - 0.5	C 0.12	0.05	-	-	-	14
Roche Cobas	11	9.3 - 14.0	P 11.65	0.3	0 - 0.5	C 0.12	0.04	-	-	-	6
Roche Cobas BILTS	12	9.1 - 13.7	P 11.42	0.18	0 - 0.6	C 0.16	0.05	-	-	-	5
Siemens Dimension	13	9.9 - 14.8	P 12.37	0.35	0 - 0.5	C 0.1	0.0	-	-	-	22
Initial Grouping bySensitivityor Principle											
Diazonium ion	14	10.0 - 15.0	P 12.48	0.74	0 - 0.5	C 0.14	0.06	-	-	-	30
Diazo/caffeine-benzoate	15	9.9 - 14.8	P 12.33	0.36	0 - 0.5	C 0.13	0.07	-	-	-	28
Other	16	10.4 - 15.6	P 12.97	0.55	0 - 0.6	C 0.16	0.07	-	-	-	7
Diazo/surfactant	17	9.2 - 13.9	P 11.55	0.28	0 - 0.5	C 0.14	0.05	-	-	-	11
Total Population											
Whole Population	18	9.9 - 14.9	P 12.42	0.79	0 - 0.5	C 0.14	0.07	-	-	-	80