



## Glycohemoglobin, 5 vial

Name	Line No.	Specimen 1			Specimen 2			Specimen 3			Specimen 4			Specimen 5			No. of Labs
		Range & Type	Mean	SD	Range & Type	Mean	SD	Range & Type	Mean	SD	Range & Type	Mean	SD	Range & Type	Mean	SD	
<b>Glycohemoglobin - HbA1c 5v</b>																	
<b>Initial Grouping by Sensitivity or Principle</b>																	
Values standardized to DCCT	1	4.4 - 6.4 C	5.44	0.24	8.2 - 11.2 P	9.7	0.33	8.3 - 11.3 P	9.8	0.39	4.4 - 6.4 C	5.4	0.24	4.5 - 6.5 C	5.52	0.17	5
Liquid chromatography	2	5.0 - 7.0 C	6.02	0.28	8.8 - 11.9 P	10.36	0.3	8.8 - 11.9 P	10.34	0.46	5.0 - 7.0 C	5.98	0.25	5.1 - 7.1 C	6.12	0.32	5
<b>Total Population</b>																	
Whole Population	3	4.7 - 6.7 C	5.67	0.36	8.6 - 11.7 P	10.15	0.58	8.6 - 11.7 P	10.15	0.55	4.6 - 6.6 C	5.64	0.34	4.7 - 6.7 C	5.71	0.38	15
<b>Glycohemoglobin - HbA1 5v</b>																	
<b>Total Population</b>																	
Whole Population	1	0 - 1.5 C	0.5	0.0	0.1 - 2.1 C	1.1	0.0	0.1 - 2.1 C	1.1	0.0	0 - 1.5 C	0.5	0.0	0 - 1.5 C	0.5	0.0	1
<b>Glycohemoglobin, Total 5v</b>																	
<b>Total Population</b>																	
Whole Population	1	10.7 - 17.9 P	14.3	0.0	10.7 - 17.9 P	14.3	0.0	10.2 - 17.0 P	13.6	0.0	11.6 - 19.3 P	15.4	0.0	11.6 - 19.4 P	15.5	0.0	1