URINE MICROALBUMIN

Name	Line No.	Specimen 1 Range & Type	Specimen 1 Mean	Specimen 1 SD	Specimen 2 Range & Type	Specimen 2 Mean	Specimen 2 SD	Specimen 3 Range & Type	Specimen 3 Mean	Specimen 3 SD	Specimen 4 Range & Type	Specimen 4 Mean	Specimen 4 SD	Specimen 5 Range & Type	Specimen 5 Mean	Specimen 5 SD	No. of Labs
Microalbumin, quantitative																	
Initial Grouping by Reagent																	
Ortho Vitros	1	29.0 - 49.0 C	39.01	0.97	72.1 - 120.1 P	96.07	2.65	-			-			-			8
Roche Cobas	2	28.5 - 48.5 C	38.49	3.26	70.4 - 117.3 P	93.86	5.33	-			-			-			10
Siemens DCA Vantage	3	25.6 - 45.6 C	35.61	2.28	69.9 - 116.4 P	93.14	4.78	-			-			-			28
Beckman Olympus	4	25.1 - 45.1 C	35.13	2.91	66.9 - 111.5 P	89.19	10.66	-			-			-			25
Roche Tina-Quant	5	29.2 - 49.2 C	39.19	1.67	69.6 - 116.0 P	92.81	3.49	-			-			-			19
Siemens Advia	6	25.5 - 45.5 C	35.5	2.32	68.0 - 113.4 P	90.7	4.4	-			-			-			4
Siemens Dimension	7	28.2 - 48.2 C	38.19	1.93	72.4 - 120.7 P	96.58	4.63	-			-			-			34
Kamiya Biomedical	8	20.0 - 40.0 C	30.0	5.89	58.4 - 97.3 P	77.8	7.1	-			-			-			5
Pointe Scientific	9	27.2 - 47.2 C	37.15	9.26	64.5 - 107.6 P	86.05	8.13	-			-			-			4
Sekisui	10	25.8 - 45.8 C	35.84	3.6	66.1 - 110.2 P	88.14	12.85	-			-			-			5
Abbott Aeroset/Architect	11	29.3 - 49.3 C	39.33	1.99	72.1 - 120.1 P	96.09	3.01	-			-			-			12
Beckman Synch LX,UniCel DxC	12	29.8 - 49.8 C	39.84	3.96	71.7 - 119.5 P	95.59	4.91	-			-			-			8
Carolina	13	21.4 - 41.4 C	31.4	1.18	63.0 - 105.0 P	84.02	3.2	-			-			-			6
Siemens DCA 2000	14	26.9 - 46.9 C	36.9	1.34	68.1 - 113.4 P	90.75	1.1	-			-			-			4
Initial Grouping by Sensitivity or Principle																	
All other methods	15	29.0 - 49.0 C	39.01	0.97	72.1 - 120.1 P	96.07	2.65	-			-			-			8
Turbidimetric	16	26.9 - 46.9 C	36.86	4.08	69.0 - 115.0 P	92.03	9.21	-			-			-			112
Siemens turbidimetric	17	25.8 - 45.8 C	35.75	2.23	69.4 - 115.7 P	92.59	4.58	-			-			-			36
Turbidity immunoinhibition	18	27.6 - 47.6 C	37.57	2.67	66.5 - 110.9 P	88.7	6.43	-			-			-			6
Chemiluminometric	19	29.3 - 49.3 C	39.33	1.99	72.1 - 120.1 P	96.09	3.01	-			-			-			12
Nephelometric	20	29.8 - 49.8 C	39.84	3.96	71.7 - 119.5 P	95.59	4.91	-			-			-			8
Total Population																	
Whole Population	21	27.0 - 47.0 C	37.01	3.71	69.4 - 115.7 P	92.56	7.91	-			-			-			183
Creatinine, quantitative																	
Initial Grouping by Reagent																	
Ortho Vitros	1	187 - 253 P	220.2	14.4	61 - 83 P	72.2	5.2	-			-			-			5
Roche Cobas	2	193 - 261 P	227.0	8.1	67 - 90 P	78.3	3.2	-			-			-			16
Siemens DCA Vantage	3	203 - 275 P	239.2	11.6	69 - 93 P	81.0	3.7	-			-			-			27
Beckman Olympus	4	193 - 261 P	226.6	8.5	65 - 87 P	75.9	2.5	-			-			-			29
Roche Tina-Quant	5	190 - 257 P	223.8	7.0	65 - 88 P	76.3	3.5	-			-			-			4
Synermed	6	195 - 264 P	229.4	19.8	66 - 90 P	78.0	7.1	-			-			-			45
Siemens Dimension	7	189 - 256 P	222.7	8.3	62 - 84 P	72.8	3.8	-			-			-			30
Beckman Coulter Synchron	8	204 - 276 P	239.6	8.1	70 - 95 P	82.2	2.3	-			-			-			5
Abbott Aeroset/Architect	9	185 - 251 P	218.0	4.4	62 - 85 P	73.5	1.1	-			-			-			13
Carolina	10	206 - 278 P	242.0	13.6	69 - 94 P	81.4	5.3	-			-			-			7
Siemens DCA 2000	11	199 - 270 P	234.5	9.3	68 - 92 P	80.3	1.5	-			-			-			4
Initial Grouping by Sensitivity or Principle																	
Jaffe/modified Jaffe	12	194 - 262 P	228.2	14.3	65 - 89 P	77.0	5.4	-			-			-			204
Total Population																	
Whole Population	13	194 - 262 P	227.8	15.0	65 - 88 P	76.8	6.0	-			-			-			209