



Glycohemoglobin, 5 vial

Name	Line No.	Specimen 1			Specimen 2			Specimen 3			Specimen 4			Specimen 5			No. of Labs
		Range & Type	Mean	SD	Range & Type	Mean	SD	Range & Type	Mean	SD	Range & Type	Mean	SD	Range & Type	Mean	SD	
Glycohemoglobin - HbA1c 5v																	
Initial Grouping by Reagent and Instrument																	
Abbott Architect & Abbott Architect c, ci, i	1	9.2 - 12.5	P 10.83	0.77	4.4 - 6.4	C 5.38	0.51	4.4 - 6.4	C 5.35	0.52	9.3 - 12.6	P 10.95	1.03	9.3 - 12.5	P 10.9	0.94	4
Initial Grouping by Reagent																	
Abbott Architect	2	9.2 - 12.5	P 10.83	0.77	4.4 - 6.4	C 5.38	0.51	4.4 - 6.4	C 5.35	0.52	9.3 - 12.6	P 10.95	1.03	9.3 - 12.5	P 10.9	0.94	4
Siemens Dimension	3	8.8 - 12.0	P 10.4	0.19	4.6 - 6.6	C 5.57	0.15	4.6 - 6.6	C 5.6	0.17	8.8 - 11.9	P 10.38	0.37	8.9 - 12.0	P 10.43	0.12	4
Initial Grouping by Sensitivity or Principle																	
Turbidity immunoinhibition	4	9.2 - 12.5	P 10.83	0.77	4.4 - 6.4	C 5.38	0.51	4.4 - 6.4	C 5.35	0.52	9.3 - 12.6	P 10.95	1.03	9.3 - 12.5	P 10.9	0.94	4
Values standardized to DCCT	5	8.8 - 11.9	P 10.34	0.21	4.5 - 6.5	C 5.5	0.2	4.5 - 6.5	C 5.48	0.29	8.8 - 11.9	P 10.34	0.34	8.8 - 12.0	P 10.4	0.12	5
Immunofluorometric-not FPIA	6	9.3 - 12.5	P 10.9	0.34	4.6 - 6.6	C 5.6	0.07	4.5 - 6.5	C 5.48	0.04	9.4 - 12.7	P 11.05	0.11	9.5 - 12.8	P 11.13	0.29	4
Total Population																	
Whole Population	7	9.1 - 12.3	P 10.73	0.58	4.6 - 6.6	C 5.61	0.45	4.6 - 6.6	C 5.58	0.47	9.2 - 12.4	P 10.81	0.66	14.0 - 19.0	P 16.49	21.67	16