



PROFICIENCY TESTING SERVICE
AMERICAN ASSOCIATION OF BIOANALYSTS
 205 West Levee St. Brownsville, TX 78520-5596
 800-234-5315 • 281-436-5357 • Fax 713-781-5008

FIRST QUADRIMESTER 2018

Alcohol and Ketones, Serum

Acetone	Specimen 1				Specimen 2				Specimen 3				Specimen 4				Specimen 5			
	Neg	Small (20 mg/dL)	Moderate (30-40 mg/dL)	Large (80-100 mg/dL)	Neg	Small (20 mg/dL)	Moderate (30-40 mg/dL)	Large (80-100 mg/dL)	Neg	Small (20 mg/dL)	Moderate (30-40 mg/dL)	Large (80-100 mg/dL)	Neg	Small (20 mg/dL)	Moderate (30-40 mg/dL)	Large (80-100 mg/dL)	Neg	Small (20 mg/dL)	Moderate (30-40 mg/dL)	Large (80-100 mg/dL)
Bayer Acetest		1	13	17	31				31				1	17	9	4	30	1		
Biorex Labs				29	29				29					5	21	3	29			
Total Population	0	3	15	50	68	0	0	0	68	0	0	0	3	28	30	7	67	1	0	0
<i>Flagging</i>	***					***	***	***		***	***	***	***					***	***	***

Correct responses are defined as those reflecting agreement among 80% or more of all participants.
 Unacceptable responses are indicated by "*****" on the Flagging line of each specimen.