



PROFICIENCY TESTING SERVICE
AMERICAN ASSOCIATION OF BIOANALYSTS
 205 West Levee St. Brownsville, TX 78520-5596
 800-234-5315 • 281-436-5357 • Fax 713-781-5008

FIRST QUADRIMESTER 2018

Pregnancy - Basic Chem

Name	Specimen 1			Specimen 2			Specimen 3			Specimen 4			Specimen 5		
	Neg	Pos	Ind	Neg	Pos	Ind	Neg	Pos	Ind	Neg	Pos	Ind	Neg	Pos	Ind
Alfa Scientific InstantView		11			11			11			11			11	
Consult hCG Combo		13			13			13			13			13	
Fisher Sure-Vue s/u hCG		14			14			14			14			14	
Quidel QuickVue OS Combo		18			18			18			18			18	
Quidel QuickVue+ OS Combo		13			13			13			13			13	
Sekisui OSOM hCG-Combo		13			13			13			13			13	
Total Population	1	146	0	1	146	0	0	147	0	1	146	0	0	147	0
<i>Flagging</i>	***		***	***		***	***		***	***		***	***		***
<i>hcg Concentration</i>	56.1 mIU/mL			62.8 mIU/mL			1809.8 mIU/mL			203.2 mIU/mL			288.0 mIU/mL		

Correct responses are defined as those reflecting agreement among 80% or more of all participants. Unacceptable responses are indicated by "*****" on the Flagging line of each specimen.

Basic Chemistry