



PROFICIENCY TESTING SERVICE
AMERICAN ASSOCIATION OF BIOANALYSTS
 205 West Levee St. Brownsville, TX 78520-5596
 800-234-5315 ♦ 281-436-5357 ♦ Fax 713-781-5008

FIRST QUADRIMESTER 2018

PREGNANCY, waived

Name	Specimen 1			Specimen 2		
	Neg	Pos	Ind	Neg	Pos	Ind
Alere	13				13	
CardinalHealth hCG Combo	23				23	
Consult Diagnostics	21				21	
Fisher Sure-Vue urine hCG	13	1		1	13	
Henry Schein OneStep	14				14	
Quidel QuickVue	20				20	
Quidel Quickvue+ OS Combo	22				22	
Sekisui OSOM hCG-Urine	18				18	
Total Population	213	2	0	1	214	0
Flagging		***	***	***		***
		<2.0 mIU/mL		2815 mIU/mL		

Correct responses are defined as those reflecting agreement among 80% or more of all participants. Unacceptable responses are indicated by "*****" on the Flagging line of each specimen.