



URINE MICROALBUMIN - CREATININE, qualitative

Microalbumin, semiquantative or qualitative

| Name | Specimen 1 | | | | | | | | | |
|-------------------------|------------|----------|----------|----------|----------|----------|----------|----------|----------|-----------|
| | Negative | Positive | 10 mg/L | 50 mg/L | 100 mg/L | 150 mg/L | 0 mg/L | 20 mg/L | 30 mg/L | 80 mg/L |
| Roche Chemstrip | | 2 | | 3 | 9 | | | | | |
| Siemens Clinitek Status | | | | | 1 | | | | 2 | 17 |
| Total Population | 0 | 2 | 0 | 3 | 9 | 1 | 0 | 0 | 4 | 22 |
| Flagging | *** | | *** | | | | *** | | | |

| Name | Specimen 2 | | | | | | | | | |
|-------------------------|------------|----------|----------|----------|----------|----------|----------|----------|-----------|----------|
| | Negative | Positive | 10 mg/L | 50 mg/L | 100 mg/L | 150 mg/L | 0 mg/L | 20 mg/L | 30 mg/L | 80 mg/L |
| Roche Chemstrip | | 2 | 2 | 8 | | | | 2 | | |
| Siemens Clinitek Status | | | 3 | | | | | | 17 | |
| Total Population | 0 | 2 | 7 | 8 | 0 | 0 | 0 | 2 | 22 | 0 |
| Flagging | *** | | | | *** | *** | *** | | | |

Creatinine, semiquantitative

| Name | Specimen 1 | | | | | Specimen 2 | | | | |
|-------------------------|------------|----------|-----------|-----------|-----------|------------|----------|-----------|-----------|-----------|
| | 10 mg/dL | 50 mg/dL | 100 mg/dL | 200 mg/dL | 300 mg/dL | 10 mg/dL | 50 mg/dL | 100 mg/dL | 200 mg/dL | 300 mg/dL |
| Siemens Clinitek Status | | 3 | 9 | 8 | | | | | 3 | 17 |
| Total Population | 0 | 3 | 13 | 11 | 0 | 0 | 0 | 0 | 5 | 22 |
| Flagging | *** | | | | *** | *** | *** | | | |

Correct responses are defined as those reflecting agreement among 80% or more of all participants. Unacceptable responses are indicated by "*****" on the Flagging line of each specimen.