Bilirubin Supplemental, 2 vial

Name	Line No.	Specimen 1 Range & Type	Specimen 1 Mean	Specimen 1 SD	Specimen 2 Range & Type	Specimen 2 Mean	Specimen 2 SD	Specimen 3 Range & Type	Specimen 3 Mean	Specimen 3 SD	Specimen 4 Range & Type	Specimen 4 Mean	Specimen 4 SD	Specimen 5 Range & Type	Specimen 5 Mean	Specimen 5 SD	No. of Labs
Bilirubin, Direct																	
Initial Grouping by Reagent and Instrument																	
Beckman Olympus & Beck Olym AU 400/600/5400	1	0 - 0.4	C 0.0	0.0	4.9 - 7.3	P 6.08	0.26	-			-			-			10
Ortho Vitros & Ortho Vitros not DT or ECi	2	0 - 0.4	C 0.0	0.0	4.3 - 6.5	P 5.39	0.39	-			-			-			13
Siemens Dimension DBI & Siemens Dimension EXL	3	0 - 0.5	C 0.05	0.05	3.4 - 5.0	P 4.19	0.2	-			-			-			35
Initial Grouping by Reagent																	
Beckman Olympus	4	0 - 0.4	C 0.01	0.02	4.8 - 7.2	P 6.04	0.28	-			-			-			18
Ortho Vitros	5	0 - 0.4	C 0.0	0.0	4.2 - 6.4	P 5.3	0.41	-			-			-			21
Roche acid diazo	6	0 - 0.5	C 0.09	0.1	3.5 - 5.3	P 4.41	0.31	-			-			-			15
Siemens Dimension DBI	7	0 - 0.5	C 0.05	0.05	3.3 - 5.0	P 4.16	0.23	-			-			-			44
Initial Grouping by Sensitivity or Principle																	
Acid diazo methods	8	0 - 0.5	C 0.06	0.06	3.6 - 5.4	P 4.53	0.75	-			-			-			109
Diazonium ion methods	9	0 - 0.4	C 0.0	0.02	4.7 - 7.0	P 5.85	0.58	-			-			-			21
Total Population																	
Whole Population	10	0 - 0.5	C 0.05	0.06	3.8 - 5.7	P 4.74	0.87	-			-			-			133
Bilirubin, Neonatal																	
Initial Grouping by Reagent																	
Ortho Vitros	1	0 - 0.4	C 0.0	0.0	8.4 - 12.6	P 10.53	0.59	-			-			-			11
Initial Grouping by Sensitivity or Principle																	
Diazonium ion	2	0 - 0.4	C 0.0	0.0	8.4 - 12.6	P 10.53	0.59	-			-			-			10
Total Population																	
Whole Population	3	0 - 0.4	C 0.03	0.05	8.2 - 12.3	P 10.21	0.61	-			-			-			22
Bilirubin, Total supplemenal																	
Initial Grouping by Reagent and Instrument																	
Siemens Dimension & Siemens Dimension EXL	1	0 - 0.5	C 0.06	0.05	7.8 - 11.8	P 9.8	0.53	-			-			-			19
Initial Grouping by Reagent																	
Ortho Vitros	2	0 - 0.5	C 0.1	0.04	8.0 - 12.0	P 10.01	0.34	-			-			-			11
Siemens Dimension	3	0 - 0.5	C 0.06	0.05	7.8 - 11.7	P 9.73	0.53	-			-			-			23
Initial Grouping by Sensitivity or Principle																	
Diazonium ion	4	0 - 0.5	C 0.1	0.04	7.7 - 11.6	P 9.66	0.85	-			-			-			21
Diazo/caffeine-benzoate	5	0 - 0.5	C 0.05	0.05	7.8 - 11.7	P 9.79	0.58	-			-			-			27
Total Population																	
Whole Population	6	0 - 0.5	C 0.08	0.06	7.8 - 11.7	P 9.72	0.82	-			-			-			61