D-Dimer

Name	Line No.	Specimen 1 Range & Type	Mean	SD	Specimen 2 Range & Type	Mean	SD	Specimen 3 Range & Type	Mean	SD	Specimen 4 Range & Type	Mean	SD	Specimen 5 Range & Type	Mean	SD	No. of Labs
D-Dimer, quant ng/mL																	
Initial Grouping by Reagent																	
Alere Triage	1	438 - 780	S 608.9	57.0	958 - 1799	S 1378.5	140.2	-			-			-			54
Siemens Stratus CS	2	459 - 635	S 547.2	29.4	1475 - 2017	S 1746.0	90.4	-			-			-			11
Initial Grouping by Sensitivity or Principle																	
low recovery	3	438 - 780	S 608.9	57.0	958 - 1799	S 1378.5	140.2	-			-			-			54
Standardized method	4	0 - 1303	S 537.7	255.1	0 - 2822	S 1362.3	486.6	-			-			-			20
Total Population																	
Whole Population	5	57 - 1173	S 614.9	186.1	444 - 2339	S 1391.9	315.8	-			-			-			84
D-Dimer, quant ugFEU/mL																	
Initial Grouping by Reagent																	
Siemens Innovance	1	1.06 - 1.66	C 1.364	0.059	2.14 - 2.93	S 2.535	0.132	-			-			-			11
Initial Grouping by Sensitivity or Principle																	
mid recovery	2	0.43 - 2.14	S 1.285	0.284	0.89 - 3.82	S 2.355	0.489	-			-			-			17
Total Population																	
Whole Population	3	0.15 - 2.21	S 1.18	0.343	0.71 - 3.79	S 2.249	0.514	-			-			-			22