



Tosoh AIA & Tosoh AIA	7	6 - 11	P 8.5	0.7	47 - 78	P 62.1	5.1	-	-	-	13
Tosoh AIA & Tosoh AIA ST	8	6 - 11	P 8.4	0.7	46 - 77	P 61.2	3.9	-	-	-	17
<b>Initial Grouping byReagent</b>											
Beckman Coulter Access	9	5 - 9	C 7.4	0.6	49 - 81	P 64.9	4.9	-	-	-	42
Roche Cobas	10	5 - 9	C 6.9	0.5	48 - 79	P 63.5	4.4	-	-	-	21
Roche Elecsys	11	5 - 9	C 7.2	0.6	48 - 79	P 63.5	4.3	-	-	-	38
Siemens Immulite	12	2 - 6	C 4.0	0.2	50 - 84	P 67.2	4.0	-	-	-	28
Siemens luminometric	13	5 - 9	C 6.9	1.0	53 - 88	P 70.2	6.5	-	-	-	13
Tosoh AIA	14	6 - 11	P 8.4	0.7	46 - 77	P 61.6	4.5	-	-	-	30
<b>Initial Grouping bySensitivityor Principle</b>											
Standardized methods	15	5 - 9	C 6.7	1.6	48 - 81	P 64.5	5.3	-	-	-	183
<b>Total Population</b>											
Whole Population	16	5 - 9	C 6.7	1.6	48 - 80	P 64.2	5.5	-	-	-	191

## Leuteinizing Hormone (LH)

<b>Initial Grouping byReagent and Instrument</b>											
Beckman Coulter Access & Beckman Coulter Access	1	1 - 5	C 3.0	0.0	42 - 70	P 55.6	3.2	-	-	-	30
Beckman Coulter Access & Beckman Coulter Dxl	2	1 - 5	C 3.3	0.4	42 - 69	P 55.4	3.2	-	-	-	11
Roche Cobas & Roche e411/e601/E170/E2010	3	2 - 6	C 4.3	0.5	64 - 107	P 85.8	5.7	-	-	-	15
Roche Elecsys & Roche e411/e601/E170/E2010	4	2 - 6	C 4.3	0.5	61 - 101	P 80.9	15.1	-	-	-	30
Siemens Immulite & Siemens Immulite 1000	5	2 - 6	C 4.0	0.0	116 - 194	P 154.8	7.9	-	-	-	23
Siemens luminometric & Siemens Centaur/Centaur CP	6	1 - 5	C 3.4	0.5	63 - 106	P 84.6	6.4	-	-	-	12
Tosoh AIA & Tosoh AIA	7	1 - 5	C 3.1	0.3	56 - 94	P 75.2	5.3	-	-	-	15
Tosoh AIA & Tosoh AIA ST	8	1 - 5	C 3.4	0.5	57 - 95	P 75.6	3.2	-	-	-	18
<b>Initial Grouping byReagent</b>											
Beckman Coulter Access	9	1 - 5	C 3.1	0.3	42 - 69	P 55.6	3.2	-	-	-	41
Roche Cobas	10	2 - 6	C 4.3	0.4	65 - 108	P 86.2	6.1	-	-	-	22
Roche Elecsys	11	2 - 6	C 4.3	0.5	61 - 101	P 81.1	13.7	-	-	-	37
Siemens Immulite	12	2 - 6	C 4.0	0.2	114 - 190	P 152.0	10.8	-	-	-	26
Siemens luminometric	13	2 - 6	C 3.5	0.5	64 - 107	P 85.3	6.7	-	-	-	13
Tosoh AIA	14	1 - 5	C 3.3	0.4	57 - 94	P 75.4	4.3	-	-	-	33
<b>Initial Grouping bySensitivityor Principle</b>											
Standardized methods	15	2 - 6	C 3.6	0.7	57 - 95	P 76.0	17.7	-	-	-	162
Siemens Immulite all	16	2 - 6	C 3.9	0.2	112 - 186	P 149.1	12.2	-	-	-	32
<b>Total Population</b>											
Whole Population	17	2 - 6	C 3.6	0.7	66 - 110	P 88.3	32.2	-	-	-	196

## Progesterone

<b>Initial Grouping byReagent and Instrument</b>											
Beckman Coulter Access & Beckman Coulter Access	1	1.5 - 2.8	P 2.15	0.28	28.4 - 52.7	P 40.57	2.96	-	-	-	31
Roche Cobas & Roche e411/e601/E170/E2010	2	1.2 - 2.2	P 1.7	0.05	31.6 - 58.7	P 45.15	1.79	-	-	-	15
Roche Elecsys & Roche e411/e601/E170/E2010	3	1.1 - 2.1	P 1.64	0.17	31.2 - 58.0	P 44.64	2.54	-	-	-	30
Siemens Immulite & Siemens Immulite 1000	4	1.3 - 2.5	P 1.9	0.19	23.3 - 43.3	P 33.34	2.27	-	-	-	24
Siemens luminometric & Siemens Centaur/Centaur CP	5	1.9 - 3.5	P 2.7	0.18	24.9 - 46.2	P 35.52	3.43	-	-	-	12
Tosoh AIA Prog II and earlier & Tosoh AIA	6	1.3 - 2.4	P 1.86	0.44	28.3 - 52.5	P 40.37	1.57	-	-	-	10
Tosoh AIA Prog II and earlier & Tosoh AIA ST	7	1.2 - 2.2	P 1.7	0.26	30.0 - 55.7	P 42.86	4.99	-	-	-	16
bioMerieux Vidas & bioMerieux mini Vidas/Vidas	8	1.3 - 2.4	P 1.82	0.47	35.4 - 65.7	P 50.53	7.62	-	-	-	10
<b>Initial Grouping byReagent</b>											
Beckman Coulter Access	9	1.5 - 2.8	P 2.12	0.27	28.0 - 52.0	P 40.03	3.65	-	-	-	38
Roche Cobas	10	1.3 - 2.3	P 1.8	0.34	31.6 - 58.8	P 45.2	1.8	-	-	-	22
Roche Elecsys	11	1.2 - 2.2	P 1.67	0.18	31.1 - 57.8	P 44.47	2.5	-	-	-	37
Siemens Immulite	12	1.3 - 2.4	P 1.88	0.19	23.1 - 42.9	P 32.99	2.57	-	-	-	27
Siemens luminometric	13	1.8 - 3.4	P 2.64	0.27	25.3 - 47.1	P 36.2	4.06	-	-	-	13
Tosoh AIA Prog II and earlier	14	1.2 - 2.3	P 1.76	0.35	29.3 - 54.5	P 41.9	4.21	-	-	-	26
Tosoh AIA Prog III	15	1.6 - 3.0	P 2.29	0.26	31.7 - 59.0	P 45.35	7.83	-	-	-	13
bioMerieux Vidas	16	1.3 - 2.4	P 1.82	0.47	35.4 - 65.7	P 50.53	7.62	-	-	-	10
<b>Initial Grouping bySensitivityor Principle</b>											
Standardized methods	17	1.4 - 2.5	P 1.94	0.41	29.9 - 55.6	P 42.77	5.43	-	-	-	164
Siemens Immulite all	18	1.3 - 2.4	P 1.86	0.24	23.1 - 42.9	P 32.98	2.68	-	-	-	35
<b>Total Population</b>											

