



Adulterated Urine

Creatinine, semiquantitative

Name	Specimen 1					Specimen 2				
	10 mg/dL	50 mg/dL	100 mg/dL	200 mg/dL	300 mg/dL	10 mg/dL	50 mg/dL	100 mg/dL	200 mg/dL	300 mg/dL
Total Population	1	0	1	0	0	2	0	0	0	0
Flagging				***	***				***	***

Creatinine, quantitative Urine Adulteration Creatinine, quantitative Urine Adulteration

Name	Line No.	Specimen 1			Specimen 2					
		Range & Type	Mean	SD	Range & Type	Mean	SD	No. of Labs		
Initial Grouping by Reagent										
Synermed	1	51 - 69	P	59.8	3.5	43533	C	5.9	0.9	12
Initial Grouping by Sensitivity or Principle										
Jaffe/modified Jaffe	2	51 - 69	P	59.6	3.4	43533	C	6.3	0.8	24
Total Population										
Whole Population	3	50 - 68	P	59.1	4.1	43533	C	6.3	0.9	26

Nitrite

Name	Specimen 1		Specimen 2	
	Neg	Positive	Neg	Positive
Total Population	2	0	0	2
Flagging		***	***	

Oxidants

Name	Specimen 1		Specimen 2	
	Neg	Positive	Neg	Positive
Total Population	5	0	0	5
Flagging		***	***	

pH

Name	Specimen 1					Specimen 2				
	<5.0	5.0	5.5	6.0	7.0	6.0	6.5	7.0	7.5	8.0
Total Population	19	0	1	0	1	2	2	14	3	0
Flagging				***	***	***				***

Specific Gravity

Name	Line No.	Specimen 1				Specimen 2				No. of Labs
		Range	Type	Mean	s.d.	Range	Type	Mean	s.d.	
Total Population										
Whole Population	1	0.998 - 1.018	C	1.0083	0.003	0.996 - 1.016	C	1.0056	0.0023	11

Adulteration Interpretation

Name	Specimen 1		Specimen 2	
	Adul	Norm	Adul	Norm
Total Population	15	4	18	1
Flagging		***		***

*Due to a lack of participant consensus, Specimen 1 was not evaluated. The intended result was Adulterated.

Correct responses are defined as those reflecting agreement among 80% or more of all participants.

Unacceptable responses are indicated by "*****" on the Flagging line of each specimen.

Ungraded results outside the optimal or intended response are indicated by "&&&" for each specimen.