



PROFICIENCY TESTING SERVICE
AMERICAN ASSOCIATION OF BIOANALYSTS
 205 West Levee St. Brownsville, TX 78520-5596
 800-234-5315 ♦ 281-436-5357 ♦ Fax 713-781-5008

FIRST QUADRIMESTER 2019

Pregnancy - Basic Chem

Name	Specimen 1			Specimen 2			Specimen 3			Specimen 4			Specimen 5		
	Neg	Pos	Ind	Neg	Pos	Ind	Neg	Pos	Ind	Neg	Pos	Ind	Neg	Pos	Ind
Alfa Scientific InstantView		14		13	1			14			14			14	
Consult hCG Combo		10		7	2	1		10			10			10	
Fisher Sure-Vue s/u hCG		11		9	2			11			11			11	
Quidel QuickVue OS Combo		16		16				16			16			16	
Total Population	2	121		108	12	3	1	122		1	122		2	121	
<i>Flagging</i>	***				***	***	***			***			***		

<i>hcg Concentration</i>	555.5 mIU/mL	2.9 mIU/mL	1006.7 mIU/mL	1665.7 mIU/mL	61.8 mIU/mL
--------------------------	--------------	------------	---------------	---------------	-------------

Correct responses are defined as those reflecting agreement among 80% or more of all participants. Unacceptable responses are indicated by "*****" on the Flagging line of each specimen.

Basic Chemistry