



## Pregnancy

Urine Name	Specimen 1			Specimen 2			Specimen 3			Specimen 4			Specimen 5		
	Neg	Pos	Ind	Neg	Pos	Ind	Neg	Pos	Ind	Neg	Pos	Ind	Neg	Pos	Ind
Alfa Scientific InstantView	11				11		11				11			11	
Consult Diagnostics	19				19		19				19		1	18	
Quidel QuickVue	21				21		21				21		1	19	1
Quidel Quickvue+ OS Combo	14				14		14				14			14	
Sekisui OSOM hCG-Urine	16				16		16				16			16	
<b>Total Population</b>	<b>173</b>	<b>0</b>	<b>0</b>	<b>0</b>	<b>173</b>	<b>0</b>	<b>172</b>	<b>1</b>	<b>0</b>	<b>2</b>	<b>171</b>	<b>0</b>	<b>6</b>	<b>165</b>	<b>2</b>
<b>Flagging</b>		***	***	***		***		***	***	***		***	***		***

Serum Name	Specimen 1			Specimen 2			Specimen 3			Specimen 4			Specimen 5		
	Neg	Pos	Ind	Neg	Pos	Ind	Neg	Pos	Ind	Neg	Pos	Ind	Neg	Pos	Ind
CardinalHealth hCG Combo	21				21		21				21			21	
<b>Total Population</b>	<b>86</b>	<b>3</b>	<b>0</b>	<b>2</b>	<b>86</b>	<b>0</b>	<b>85</b>	<b>3</b>	<b>0</b>	<b>2</b>	<b>86</b>		<b>2</b>	<b>86</b>	<b>0</b>
<b>Flagging</b>		***	***	***		***		***	***	***			***		***

*hcg Concentration*                      <2.0 mIU/mL                      55.0 mIU/mL                      <2.0 mIU/mL                      55.0 mIU/mL                      55.0 mIU/mL

Correct responses are defined as those reflecting agreement among 80% or more of all participants.  
 Unacceptable responses are indicated by "\*\*\*\*\*" on the Flagging line of each specimen.