



PROFICIENCY TESTING SERVICE
AMERICAN ASSOCIATION OF BIOANALYSTS
 205 West Levee St. Brownsville, TX 78520-5596
 800-234-5315 ♦ 281-436-5357 ♦ Fax 713-781-5008

FIRST QUADRIMESTER 2019

PREGNANCY, waived

Name	Specimen 1			Specimen 2			
	Neg	Pos	Ind	Neg	Pos	Ind	
CardinalHealth hCG Combo	21				21		
Consult Diagnostics	30				30		
Fisher Sure-Vue urine hCG	23				23		
Medline	10				10		
OneStep 3rd party, OEM waiv	10				10		
Quidel QuickVue	15	1		1	15		
Sekisui OSOM hCG-Urine	19				19		
Siemens Clinitest	14				14		
Total Population	197	1	0	1	197	0	
Flagging		***	***	***		***	
		<2.0 mIU/mL			55.0 mIU/mL		

Correct responses are defined as those reflecting agreement among 80% or more of all participants. Unacceptable responses are indicated by "*****" on the Flagging line of each specimen.