



## Specific Gravity - Urinalysis

Name	Line No.	Range	Type	Mean	s.d.	No. of Labs
<b>Specific Gravity - Urinalysis</b>						
<b>Initial Grouping by Method</b>						
Siemens Clinitek Status, Plus	1	1.009 - 1.029	C	1.0192	0.0026	276
Siemens Clinitek 100	2	1.009 - 1.029	C	1.0194	0.0028	9
Siemens Reagent Strips	3	1.002 - 1.022	C	1.0124	0.0037	77
Roche Chemstrip 10 SG	4	0.995 - 1.015	C	1.005	0.0017	17
Stanbio	5	1.003 - 1.023	C	1.0133	0.0024	3
Mckesson MediLab 10SG	6	1.009 - 1.029	C	1.0193	0.0037	14
Siemens Clinitek Advantus	7	1.007 - 1.027	C	1.0174	0.0037	34
Siemens Clinitek 10	8	1.005 - 1.025	C	1.0146	0.0046	26
Roche Chemstrip 10 MD	9	0.997 - 1.017	C	1.0067	0.0041	9
Mckesson 120	10	1.009 - 1.029	C	1.0187	0.0034	15
PSS Consult	11	1.006 - 1.026	C	1.0157	0.0022	23
YPC Uriscan	12	1.004 - 1.024	C	1.0141	0.0047	11
All refractometers	13	1.003 - 1.023	C	1.0125	0.0011	16
Medline	14	1.009 - 1.029	C	1.019	0.0037	5
NDC Pro Advantage	15	1.01 - 1.03	C	1.02	0	4
Germaine AimStick	16	1.009 - 1.029	C	1.0187	0.0041	4
YD Diagnostics	17	1.006 - 1.026	C	1.0157	0.0017	7
Siemens Clinitek 50	18	1.005 - 1.025	C	1.0146	0.0052	12
Iris iChem	19	1.002 - 1.022	C	1.012	0.0006	5
Roche Urisys 1100	20	1.001 - 1.021	C	1.011	0.0037	10
Siemens Clinitek 500	21	1.009 - 1.029	C	1.0189	0.0038	19
Roche Cobas U 411	22	1.001 - 1.021	C	1.0113	0.0022	8
Roche Chemstrip 10 UA	23	0.997 - 1.017	C	1.0067	0.0024	3
RAC Clarity	24	1.013 - 1.033	C	1.0233	0.0047	3
Jant Pharmacal Accutest	25	1.0 - 1.02	C	1.01	0	3
All automated chem analyzer	26	1.006 - 1.026	C	1.016	0.0029	3
<b>Initial Grouping by Sensitivity or Principle</b>						
All strips	27	1.007 - 1.027	C	1.017	0.0045	580
All direct physical methods	28	1.002 - 1.022	C	1.0124	0.0013	19
Other	29	1.009 - 1.029	C	1.0193	0.0062	7
Other; uncertain	30	1.006 - 1.026	C	1.016	0.0029	3
<b>Total Population</b>						
Whole Population	31	1.007 - 1.027	C	1.0165	0.0049	646

Correct responses are defined as those reflecting agreement among 80% or more of all participants.  
 Unacceptable responses are indicated by "\*\*\*\*\*" on the Flagging line of each specimen.