



URINE MICROALBUMIN - CREATININE, qualitative

Microalbumin, semiquantitative or qualitative

Name	Specimen 1									
	Negative	Positive	10 mg/L	50 mg/L	100 mg/L	150 mg/L	0 mg/L	20 mg/L	30 mg/L	80 mg/L
Roche Chemstrip		1			8			1		1
Siemens Clinitek Status									7	15
Total Population	0	1	1	0	8	0	0	1	8	20
Flagging	***		***				***			

Name	Specimen 2									
	Negative	Positive	10 mg/L	50 mg/L	100 mg/L	150 mg/L	0 mg/L	20 mg/L	30 mg/L	80 mg/L
Roche Chemstrip		1		7				3		
Siemens Clinitek Status			10						12	
Total Population	0	1	11	7	0	0	0	3	17	0
Flagging	***				***	***	***			

Creatinine, semiquantitative

Name	Specimen 1					Specimen 2				
	10 mg/dL	50 mg/dL	100 mg/dL	200 mg/dL	300 mg/dL	10 mg/dL	50 mg/dL	100 mg/dL	200 mg/dL	300 mg/dL
Siemens Clinitek Status				1	22				1	22
Total Population	0	0	0	2	27	0	0	0	1	28
Flagging	***	***				***	***			

Correct responses are defined as those reflecting agreement among 80% or more of all participants. Unacceptable responses are indicated by "*****" on the Flagging line of each specimen.