Ammonia,Serum

		Specimen 1			Specimen 2			Specimen 3			Specimen 4			Specimen 5			
Name	**Line No.**	**Range & Type**	**Mean**	**SD**	**Range & Type**	**Mean**	**SD**	**Range & Type**	**Mean**	**SD**	**Range & Type**	**Mean**	**SD**	**Range & Type**	**Mean**	**SD**	**No. of Labs**
Ammonia, Serum																	
Initial Grouping by Reagent and Instrument																	
Siemens Dimension & Siemens Dimension EXL	1	144 - 176	P 160.3	5.8	69 - 89	C 79.3	3.8	-			-			-			32
Initial Grouping by Reagent																	
Ortho Vitros	2	120 - 147	P 133.7	6.2	48 - 68	C 58.2	5.6	-			-			-			11
Siemens Dimension	3	144 - 176	P 160.2	5.9	69 - 89	C 79.2	3.7	-			-			-			36
Initial Grouping by Sensitivity or Principle																	
All other methods	4	143 - 174	P 158.5	12.1	69 - 89	C 79.1	5.9	-			-			-			54
Dry Slide or Cartridge	5	120 - 147	P 133.7	6.2	48 - 68	C 58.2	5.6	-			-			-			11
Total Population																	
Whole Population	6	139 - 170	P 154.2	14.6	66 - 86	C 75.5	9.8	-			-			-			65