



AMERICAN ASSOCIATION OF BIOANALYSTS
PROFICIENCY TESTING
11931 Wickchester Ln., Ste 200 - Houston, TX 77043
 800-234-5315 - 281-436-5357 - Fax 713-781-5008

FIRST QUADRIMESTER 2022

Pregnancy - Basic Chem

Name	Specimen 1			Specimen 2			Specimen 3			Specimen 4			Specimen 5		
	Neg	Pos	Ind	Neg	Pos	Ind	Neg	Pos	Ind	Neg	Pos	Ind	Neg	Pos	Ind
Alfa Scientific InstantView		8			8		8				8				8
CardinalHealth hCG Combo		4			4		3	1			4				4
Consult hCG Combo		7			7		6	1			7				7
Fisher Sure-Vue hCG-STAT		5			5		5				5				5
Fisher Sure-Vue s/u hCG		7			7		6	1			7				7
Quidel QuickVue OS Combo		5			5		5				5				5
Total Population	0	81	1	0	82	0	71	11	0	0	82	0	1	81	0
Flagging	***		***	***		***		***	***	***		***	***	***	***

hcg Concentration 59.4 mIU/mL 1810.9 mIU/mL 0.8 mIU/mL 1064.7 mIU/mL 501.1 mIU/mL

Correct responses are defined as those reflecting agreement among 80% or more of all participants.
 Unacceptable responses are indicated by "*****" on the Flagging line of each specimen.

Basic Chemistry