AMERICAN ASSOCIATION OF BIOANALYSTS
PROFICIENCY TESTING
11931 Wickchester Ln., Ste 200 - Houston, TX 77043
800-234-5315 - 281-436-5357 - Fax 713-781-5008

FIRST QUADRIMESTER 2022

Pregnancy

Urine

	Specimen 1			Specimen 2			Specimen 3			Specimen 4			Specimen 5		
Name	**Neg**	**Pos**	**Ind**	**Neg**	**Pos**	**Ind**	**Neg**	**Pos**	**Ind**	**Neg**	**Pos**	**Ind**	**Neg**	**Pos**	**Ind**
Alere	1	3		3			4				4			4	
Alfa Scientific InstantView		6		5			5				5			5	
Beckman ICON 20/25 s/u		4		4			4				4			4	
CardinalHealth hCG Combo		6		6			6				6			6	
Consult Diagnostics		16		16			16				16			16	
Fisher Sure-Vue hCG-STAT		4		4			4				4			4	
Fisher Sure-Vue urine hCG		4		4			4				4			4	
Henry Schein One Step +		8		8			8				8			8	
McKesson	1	10		11			11				11			11	
Quidel QuickVue		6		6			6				6			6	
Sekisui OSOM hCG-Combo		5		5			5				5			5	
Total Population	**3**	**99**	**0**	**100**	**0**	**0**	**101**	**0**	**0**	**0**	**101**	**0**	**0**	**101**	**0**
Flagging	***		***		***	***		***	***	***		***	***		***

Serum

	Specimen 1			Specimen 2			Specimen 3			Specimen 4			Specimen 5		
Name	**Neg**	**Pos**	**Ind**	**Neg**	**Pos**	**Ind**	**Neg**	**Pos**	**Ind**	**Neg**	**Pos**	**Ind**	**Neg**	**Pos**	**Ind**
Alere		4		3	1		4				4			4	
Alfa Scientific InstantView		4		4			4				4			4	
CardinalHealth hCG Combo		7		7			7				7			7	
Consult Diagnostics		7		7			7				7			7	
Fisher Sure-Vue s/u hCG		5		5			5				5			5	
Total Population	**1**	**48**	**0**	**47**	**1**	**0**	**48**	**0**	**0**	**1**	**47**	**0**	**0**	**48**	**0**
Flagging	***		***		***	***		***	***	***		***	***		***

hcg Concentration	56.1 mIU/mL	<2.0 mIU/mL	<2.0 mIU/mL	56.1 mIU/mL	56.1 mIU/mL

Correct responses are defined as those reflecting agreement among 80% or more of all participants.
Unacceptable responses are indicated by "*****" on the Flagging line of each specimen.