



p2PSA

Name	Line No.	Specimen 1					Specimen 2					No. of Labs
		Range & Type	Mean	SD	CV	Range & Type	Mean	SD	CV			
p2PSA												
Last Updated: March 1st, 2022 15:42												
Initial Grouping by Reagent												
Beckman Coulter DxI	1	7.0 - 13.1 P	10.06	0.26	2.6%	72.0 - 133.6 P	102.8	6.09	5.9%	5		
Beckman Coulter Access	2	7.3 - 13.5 P	10.4	0.66	6.3%	68.9 - 127.9 P	98.42	2.38	2.4%	5		
Initial Grouping by Sensitivity or Principle												
Beckman Coulter, all	3	7.2 - 13.3 P	10.23	0.53	5.2%	70.4 - 130.8 P	100.61	5.12	5.1%	10		
Total Population												
Whole Population	4	7.2 - 13.3 P	10.23	0.53	5.2%	70.4 - 130.8 P	100.61	5.12	5.1%	10		