



**PROFICIENCY TESTING SERVICE
AMERICAN ASSOCIATION OF BIOANALYSTS**

205 West Levee St. ♦ Brownsville, TX 78520-5596
800-234-5315 ♦ 281-436-5357 ♦ Fax 713-781-5008

SECOND QUADRIMESTER 2011

Acetone

Acetone	Specimen 1				Specimen 2				Specimen 3				Specimen 4				Specimen 5			
	Neg	Small (20 mg/dL)	Moderate (30-40 mg/dL)	Large (80-100 mg/dL)	Neg	Small (20 mg/dL)	Moderate (30-40 mg/dL)	Large (80-100 mg/dL)	Neg	Small (20 mg/dL)	Moderate (30-40 mg/dL)	Large (80-100 mg/dL)	Neg	Small (20 mg/dL)	Moderate (30-40 mg/dL)	Large (80-100 mg/dL)	Neg	Small (20 mg/dL)	Moderate (30-40 mg/dL)	Large (80-100 mg/dL)
Bayer Acetest	78					60	17	1		4	35	39	78				78			
Gas chromat	1					0	1			1	1		1				1			
Total Population	95	0	0	0	0	70	24	1	0	5	42	48	95	0	0	0	95	0	0	0
<i>Flagging</i>		*****	*****	*****	*****				*****					*****	*****	*****		*****	*****	*****

Correct responses are defined as those reflecting agreement among 80% or more of all participants.
Unacceptable responses are indicated by "*****" on the Flagging line of each specimen.