D-Dimer, quantitative D-Dimer, quantitative

Name	Line No.	Specimen 1 Range & Type		Mean	SD	Specimen 2 Range & Type		Mean	SD	Specimen 3 Range & Type	Mean	SD	Specimen 4 Range & Type	Mean	SD	Specimen 5 Range & Type	Mean	SD	No. of Labs
D-Dimer, quantitative																			
Initial Grouping by Method																			
Biosite Triage	1	102.0 - 591.0	S	346.4	81.6	941.0 - 2234.0	S	1587.2	215.5	-			-			-			117
Siemens Stratus CS	2	365.0 - 526.0	S	445.6	26.7	2211.0 - 8995.0	S	5603.3	1130.7	-			-			-			28
Siemens Sysmex Advanced	3	1082.0 - 1791.0	S	1436.6	118.2	1878.0 - 2571.0	S	2224.3	115.5	-			-			-			17
Initial Grouping by Sensitivity or Principle																			
low recovery	4	102.0 - 591.0	S	346.4	81.6	941.0 - 2234.0	S	1587.2	215.5	-			-			-			117
mid recovery	5	1082.0 - 1791.0	S	1436.6	118.2	1878.0 - 2571.0	S	2224.3	115.5	-			-			-			17
high recovery	6	0.0 - 1163.0	S	517.7	215.2	0.0 - 10954.0	S	4091.9	2287.5	-			-			-			43
IL Non ACL 7000	7	327.0 - 610.0	S	468.6	47.3	506.0 - 844.0	S	675.2	56.2	-			-			-			12
Total Population																			
Whole Population	8	0.0 - 1450.0	S	481.0	322.8	0.0 - 6677.0	S	2054.7	1540.8	-			-			-			209