



**PROFICIENCY TESTING SERVICE
AMERICAN ASSOCIATION OF BIOANALYSTS**

205 West Levee St. ♦ Brownsville, TX 78520-5596
800-234-5315 ♦ 281-436-5357 ♦ Fax 713-781-5008

PARTICIPANT STATISTICS

SECOND QUADRIMESTER 2011

D-Dimer, Qualitative

Name
American Diag. Simplify
bioMerieux Vidas D-Dimer
Biosite Triage
Diagnostica Stago Liatest
IL ACL Futura/Advance
Inverness Clearview
Pacific Hemostasis
Siemens Dimertest
Total Population
Flagging

Specimen 1	
Positive	Negative
1	
1	
	3
2	
1	
7	12
1	1
1	1
14	17

Specimen 2	
Positive	Negative
1	
1	
3	
2	
1	
19	
1	1
2	
30	1

Due to a lack of participant consensus, Specimen 1 was not evaluated this event.

Correct responses are defined as those reflecting agreement among 80% or more of all participants. Unacceptable responses are indicated by "*****" on the Flagging line of each specimen.