



High Sensitivity C-Reactive Protein

Name	Line No.	Specimen 1			Specimen 2			No. of Labs		
		Range & Type	Mean	SD	Range & Type	Mean	SD			
High Sensitivity C-Reactive Protein										
Initial Grouping by Method										
Beckman Synchron turbidimet	1	1.02 - 1.49	S	1.256	0.079	0.22 - 10.79	S	5.508	1.762	11
Beckman Unicel	2	0.95 - 1.43	S	1.188	0.079	4.94 - 7.02	S	5.980	0.348	13
Carolina immunoturbid	3	0.62 - 1.83	S	1.225	0.201	4.95 - 8.02	S	6.482	0.512	13
Roche COBAS Integra	4	0.0 - 8.76	S	1.814	2.315	2.41 - 7.19	S	4.804	0.797	11
Siemens CCRP	5	0.89 - 2.56	S	1.726	0.277	4.58 - 10.05	S	7.315	0.912	16
Siemens Extended RCRP	6	1.04 - 2.74	S	1.889	0.284	5.38 - 8.27	S	6.822	0.482	11
Siemens Immulite 2000	7	0.0 - 12.77	S	2.356	3.473	4.27 - 7.56	S	5.917	0.548	10
Initial Grouping by Sensitivity or Principle										
All wide range methods	8	0.0 - 2.75	S	1.361	0.462	0.68 - 10.81	S	5.746	1.690	17
All Roche HS methods	9	0.0 - 6.91	S	1.653	1.753	1.61 - 7.85	S	4.730	1.040	22
All direct quantitative meth	10	0.0 - 5.24	S	1.340	1.301	0.0 - 26.29	S	6.658	6.544	106
All quantitations by dilutn	11	0.0 - 6.89	S	1.706	1.728	3.62 - 9.87	S	6.746	1.040	39
Total Population										
Whole Population	12	0.66 - 1.91	S	1.283	0.208	3.77 - 8.34	S	6.055	0.763	190