



**PROFICIENCY TESTING SERVICE
AMERICAN ASSOCIATION OF BIOANALYSTS**

205 West Levee St. ♦ Brownsville, TX 78520-5596
800-234-5315 ♦ 281-436-5357 ♦ Fax 713-781-5008

PARTICIPANT STATISTICS

SECOND QUADRIMESTER 2016

CARDIAC MARKERS / ISOENZYMES

	Specimen 1		Specimen 2		Specimen 3		Specimen 4		Specimen 5	
	Unelevated	Elevated	Unelevated	Elevated	Unelevated	Elevated	Unelevated	Elevated	Unelevated	Elevated
D-Dimer, qualitative										
Alere Clearview	3		1	2	3		1	2		3
Total Population	4	0	1	3	3	0	1	2	0	3
Flagging		***				***			***	

*Due to a lack of participant consensus, Specimens 2 and 4 were not evaluated. Please compare your results to the Quantitative results for evaluation.

	Specimen 1		Specimen 2		Specimen 3		Specimen 4		Specimen 5	
	Unelevated	Elevated	Unelevated	Elevated	Unelevated	Elevated	Unelevated	Elevated	Unelevated	Elevated
Troponin I										
Total Population	0	1	0	1	1	0	0	1	0	1
Flagging	***		***			***	***		***	

Please refer to the Quantitative results for evaluation.

Correct responses are defined as those reflecting agreement among 80% or more of all participants. Unacceptable responses are indicated by "*****" on the Flagging line of each specimen.