

| | | | | | | | | | | | |
|---|----|---------|--------|------|-------------|---------|------|---|---|---|----|
| Diazo/surfactant | 7 | 0 - 0.6 | C 0.18 | 0.04 | 16.0 - 24.0 | P 20.0 | 0.58 | - | - | - | 4 |
| Total Population | | | | | | | | | | | |
| Whole Population | 8 | 0 - 0.5 | C 0.07 | 0.07 | 14.8 - 22.2 | P 18.54 | 2.16 | - | - | - | 34 |
| Bilirubin, Total supplemental | | | | | | | | | | | |
| Initial Grouping by Reagent and Instrument | | | | | | | | | | | |
| Abbott Architect & Abbott Architect c, ci, i | 1 | 0 - 0.5 | C 0.11 | 0.03 | 14.9 - 22.3 | P 18.6 | 0.43 | - | - | - | 8 |
| Beckman Coulter & Beck Coult Unicel DXC | 2 | 0 - 0.6 | C 0.23 | 0.05 | 15.3 - 22.9 | P 19.1 | 0.29 | - | - | - | 5 |
| Beckman Olympus & Beck Olym AU 400/600/5400 | 3 | 0 - 0.6 | C 0.15 | 0.05 | 15.8 - 23.7 | P 19.78 | 0.5 | - | - | - | 4 |
| Ortho Vitros & Ortho Vitros 3600, 5600 | 4 | 0 - 0.6 | C 0.18 | 0.13 | 16.1 - 24.2 | P 20.13 | 0.75 | - | - | - | 4 |
| Ortho Vitros & Ortho Vitros not DT or ECi | 5 | 0 - 0.5 | C 0.13 | 0.05 | 16.5 - 24.7 | P 20.6 | 0.58 | - | - | - | 7 |
| Siemens Dimension & Siemens Dimension EXL | 6 | 0 - 0.5 | C 0.12 | 0.04 | 15.5 - 23.3 | P 19.41 | 0.81 | - | - | - | 18 |
| Siemens Dimension & Siemens Dimension Xpand | 7 | 0 - 0.5 | C 0.13 | 0.04 | 15.3 - 23.0 | P 19.15 | 0.26 | - | - | - | 4 |
| Initial Grouping by Reagent | | | | | | | | | | | |
| Abbott Architect | 8 | 0 - 0.5 | C 0.11 | 0.03 | 14.9 - 22.3 | P 18.6 | 0.43 | - | - | - | 8 |
| Beckman Coulter | 9 | 0 - 0.6 | C 0.22 | 0.04 | 15.4 - 23.0 | P 19.2 | 0.31 | - | - | - | 6 |
| Beckman Olympus | 10 | 0 - 0.6 | C 0.16 | 0.05 | 15.9 - 23.9 | P 19.92 | 0.53 | - | - | - | 5 |
| Ortho Vitros | 11 | 0 - 0.6 | C 0.15 | 0.09 | 16.2 - 24.3 | P 20.29 | 0.65 | - | - | - | 14 |
| Roche Cobas | 12 | 0 - 0.5 | C 0.12 | 0.04 | 15.7 - 23.6 | P 19.64 | 1.04 | - | - | - | 6 |
| Roche Cobas BLTS | 13 | 0 - 0.6 | C 0.15 | 0.05 | 14.9 - 22.4 | P 18.63 | 0.16 | - | - | - | 4 |
| Siemens Dimension | 14 | 0 - 0.5 | C 0.12 | 0.04 | 15.5 - 23.2 | P 19.34 | 0.72 | - | - | - | 24 |
| Initial Grouping by Sensitivity or Principle | | | | | | | | | | | |
| Diazonium ion | 15 | 0 - 0.5 | C 0.14 | 0.08 | 15.8 - 23.7 | P 19.75 | 0.89 | - | - | - | 30 |
| Diazo/caffeine-benzoate | 16 | 0 - 0.5 | C 0.13 | 0.05 | 15.5 - 23.2 | P 19.32 | 0.67 | - | - | - | 30 |
| Other | 17 | 0 - 0.5 | C 0.14 | 0.07 | 16.0 - 24.0 | P 19.97 | 0.68 | - | - | - | 8 |
| Diazo/surfactant | 18 | 0 - 0.5 | C 0.13 | 0.05 | 15.4 - 23.0 | P 19.19 | 0.93 | - | - | - | 10 |
| Total Population | | | | | | | | | | | |
| Whole Population | 19 | 0 - 0.5 | C 0.14 | 0.07 | 15.7 - 23.5 | P 19.57 | 1.06 | - | - | - | 83 |