



Glycohemoglobin, 5 vial

Name	Line No.	Specimen 1			Specimen 2			Specimen 3			Specimen 4			Specimen 5			No. of Labs
		Range & Type	Mean	SD	Range & Type	Mean	SD	Range & Type	Mean	SD	Range & Type	Mean	SD	Range & Type	Mean	SD	
Glycohemoglobin - HbA1c 5v																	
Initial Grouping by Reagent																	
Beckman Olympus	1	7.2 - 9.7	P 8.42	0.35	4.6 - 6.6	C 5.63	0.13	7.1 - 9.6	P 8.32	0.4	4.6 - 6.6	C 5.63	0.19	7.0 - 9.5	P 8.28	0.31	4
Initial Grouping by Sensitivity or Principle																	
Values standardized to DCCT	2	7.1 - 9.7	P 8.4	0.29	4.6 - 6.6	C 5.6	0.12	7.1 - 9.6	P 8.33	0.32	4.6 - 6.6	C 5.62	0.17	7.1 - 9.5	P 8.3	0.26	6
Liquid chromatography	3	7.1 - 9.6	P 8.32	1.0	5.0 - 7.0	C 5.98	0.48	7.2 - 9.8	P 8.5	0.72	4.9 - 6.9	C 5.88	0.55	7.3 - 9.9	P 8.6	0.58	4
Total Population																	
Whole Population	4	7.2 - 9.8	P 8.51	0.62	4.8 - 6.8	C 5.77	0.32	7.2 - 9.8	P 8.51	0.5	4.7 - 6.7	C 5.73	0.34	7.3 - 9.9	P 8.59	0.46	14
Glycohemoglobin - HbA1 5v																	
Total Population																	
Whole Population	1	0 - 1.9	C 0.9	0.0	0 - 1.4	C 0.4	0.0	0 - 1.9	C 0.9	0.0	0 - 1.5	C 0.5	0.0	0 - 1.9	C 0.9	0.0	1
Glycohemoglobin, Total 5v																	
Total Population																	
Whole Population	1	9.5 - 15.8	P 12.6	0.0	8.6 - 14.3	P 11.4	0.0	10.1 - 16.9	P 13.5	0.0	10.0 - 16.6	P 13.3	0.0	10.1 - 16.9	P 13.5	0.0	1