



**PROFICIENCY TESTING SERVICE
AMERICAN ASSOCIATION OF BIOANALYSTS**

205 West Levee St. ♦ Brownsville, TX 78520-5596
800-234-5315 ♦ 281-436-5357 ♦ Fax 713-781-5008

PARTICIPANT STATISTICS

SECOND QUADRIMESTER 2017

CARDIAC MARKERS / ISOENZYMES

	Specimen 1		Specimen 2		Specimen 3		Specimen 4		Specimen 5	
	Unelevated	Elevated	Unelevated	Elevated	Unelevated	Elevated	Unelevated	Elevated	Unelevated	Elevated
Troponin I										
Total Population	0	1	1	0	0	1	0	1	0	1
Flagging	***			***	***		***		***	

*Please evaluate vs. Tnl quantitative stats.

Correct responses are defined as those reflecting agreement among 80% or more of all participants. Unacceptable responses are indicated by "*****" on the Flagging line of each specimen.