



PROFICIENCY TESTING SERVICE
 AMERICAN ASSOCIATION OF BIOANALYSTS
 205 West Levee St. Brownsville, TX 78520-5596
 800-234-5315 • 281-436-5357 • Fax 713-781-5008

SECOND QUADRIMESTER 2018

Alcohol and Ketones, Serum

Acetone	Specimen 1				Specimen 2				Specimen 3				Specimen 4				Specimen 5			
	Neg	Small (20 mg/dL)	Moderate (30-40 mg/dL)	Large (80-100 mg/dL)	Neg	Small (20 mg/dL)	Moderate (30-40 mg/dL)	Large (80-100 mg/dL)	Neg	Small (20 mg/dL)	Moderate (30-40 mg/dL)	Large (80-100 mg/dL)	Neg	Small (20 mg/dL)	Moderate (30-40 mg/dL)	Large (80-100 mg/dL)	Neg	Small (20 mg/dL)	Moderate (30-40 mg/dL)	Large (80-100 mg/dL)
Bayer Acetest	29				6	9	14		29				29					7	5	17
Biorex Labs	30					4	26		30				30						4	26
Germaine Aim	11	1			1	8	1	2	11	1			12					4	6	2
Total Population	70	1	0	0	1	14	14	42	70	1	0	0	71	0	0	0	0	11	15	45
<i>Flagging</i>		***	***	***	***					***	***	***		***	***	***	***			

Correct responses are defined as those reflecting agreement among 80% or more of all participants.
 Unacceptable responses are indicated by "*****" on the Flagging line of each specimen.