



PROFICIENCY TESTING SERVICE
AMERICAN ASSOCIATION OF BIOANALYSTS
 205 West Levee St. Brownsville, TX 78520-5596
 800-234-5315 • 281-436-5357 • Fax 713-781-5008

SECOND QUADRIMESTER 2018

Pregnancy - Basic Chem

Name	Specimen 1			Specimen 2			Specimen 3			Specimen 4			Specimen 5		
	Neg	Pos	Ind	Neg	Pos	Ind	Neg	Pos	Ind	Neg	Pos	Ind	Neg	Pos	Ind
Alfa Scientific InstantView		14			14			14			14			14	
Consult hCG Combo		10			10			10			10			10	
Fisher Sure-Vue s/u hCG		12			12			12			12			12	
Quidel QuickVue OS Combo		17			17			17			17			17	
Quidel QuickVue+ OS Combo		15			15			15			15			15	
Sekisui OSOM hCG-Combo		12			12			12			12			12	
Total Population	0	142	0	0	142	0	0	142	0	0	142	0	0	142	0
Flagging	***		***	***		***	***		***	***		***	***		***
<i>hcg Concentration</i>	930.5 mIU/mL			430.7 mIU/mL			1,430.3 mIU/mL			220.5 mIU/mL			297.2 mIU/mL		

Correct responses are defined as those reflecting agreement among 80% or more of all participants. Unacceptable responses are indicated by "*****" on the Flagging line of each specimen.

Basic Chemistry