



**PROFICIENCY TESTING SERVICE**  
**AMERICAN ASSOCIATION OF BIOANALYSTS**  
 205 West Levee St. Brownsville, TX 78520-5596  
 800-234-5315 • 281-436-5357 • Fax 713-781-5008

SECOND QUADRIMESTER 2018

## Fecal Occult Blood

Fecal Occult Blood	Specimen 1		Specimen 2	
	Negative	Positive	Negative	Positive
SENSA	1	71	71	1
Beck/Coult Hemmocult	1	62	59	4
Consult Diagnostics		11	11	
Enterix InSure		24	24	
Fischer Sure-View FOB		12	12	
Helena ColoScreen		13	13	
Hemosure	1	54	54	1
Immunost Hema Screen		10	10	
Polymedco OC-Light		10	10	
Propper Seracut		17	17	
Quidel IFOB		10	10	
<b>Total Population</b>	<b>4</b>	<b>318</b>	<b>315</b>	<b>7</b>
<b>Flagging</b>	***			***

Correct responses are defined as those reflecting agreement among 80% or more of all participants  
 Unacceptable responses are indicated by "\*\*\*\*\*" on the Flagging line of each specimen