



PROFICIENCY TESTING SERVICE  
 AMERICAN ASSOCIATION OF BIOANALYSTS  
 205 West Levee St. Brownsville, TX 78520-5596  
 800-234-5315 • 281-436-5357 • Fax 713-781-5008

SECOND QUADRIMESTER 2018

## Gastric Occult Blood

### Gastric Occult Blood

Gastric Occult Blood	Specimen 1		Specimen 2	
	Neg	Pos	Neg	Pos
Beckman Coulter Gastrocult	11			11
<b>Total Population</b>	<b>12</b>	<b>0</b>	<b>0</b>	<b>12</b>
<i>Flagging</i>		***	***	

pH - Gastric	Specimen 1					Specimen 2				
	2.0	3.0	4.0	5.0	>5.0	2.0	3.0	4.0	5.0	>5.0
<b>Total Population</b>	<b>0</b>	<b>3</b>	<b>6</b>	<b>0</b>	<b>0</b>	<b>2</b>	<b>6</b>	<b>1</b>	<b>0</b>	<b>0</b>
<i>Flagging</i>	***				***				***	***

Correct responses are defined as those reflecting agreement among 80% or more of all participants. Unacceptable responses are indicated by "\*\*\*\*\*" on the Flagging line of each specimen.