



PROFICIENCY TESTING SERVICE
AMERICAN ASSOCIATION OF BIOANALYSTS
 205 West Levee St. Brownsville, TX 78520-5596
 800-234-5315 • 281-436-5357 • Fax 713-781-5008

SECOND QUADRIMESTER 2018

PREGNANCY, waived

Name	Specimen 1			Specimen 2		
	Neg	Pos	Ind	Neg	Pos	Ind
Alere	1	12		13		
CardinalHealth hCG Combo		22		21	1	
Consult Diagnostics		25		25		
Fisher Sure-Vue urine hCG		17		17		
Henry Schein OneStep		12		12		
McKesson		10		10		
Quidel QuickVue		19		19		
Quidel Quickvue+ OS Combo		21		21		
Sekisui OSOM hCG-Urine		20		20		
Siemens Clinitest		12		12		
Total Population	1	224		224	1	
Flagging	***				***	

2815 mIU/mL

<2.0 mIU/mL

Correct responses are defined as those reflecting agreement among 80% or more of all participants.
 Unacceptable responses are indicated by "*****" on the Flagging line of each specimen.