



URINE MICROALBUMIN - CREATININE, qualitative

Microalbumin, semiquantitative or qualitative

Name	Specimen 1									
	Negative	Positive	10 mg/L	50 mg/L	100 mg/L	150 mg/L	0 mg/L	20 mg/L	30 mg/L	80 mg/L
Roche Chemstrip		2			10					
Siemens Clinitek Status						8				13
Total Population	0	2	0	0	10	8	0	0	0	20
Flagging	***		***				***			

Name	Specimen 2									
	Negative	Positive	10 mg/L	50 mg/L	100 mg/L	150 mg/L	0 mg/L	20 mg/L	30 mg/L	80 mg/L
Roche Chemstrip		2		9				1		
Siemens Clinitek Status	2		7						11	1
Total Population	2	2	12	9	0	0	0	1	13	1
Flagging	***					***	***			

Creatinine, semiquantitative

Name	Specimen 1					Specimen 2				
	10 mg/dL	50 mg/dL	100 mg/dL	200 mg/dL	300 mg/dL	10 mg/dL	50 mg/dL	100 mg/dL	200 mg/dL	300 mg/dL
Siemens Clinitek Status		2	15	5					3	19
Total Population	0	2	19	7	1	0	0	0	6	23
Flagging	***				***	***	***			

Correct responses are defined as those reflecting agreement among 80% or more of all participants. Unacceptable responses are indicated by "*****" on the Flagging line of each specimen.