Ammonia,Serum

		Specimen 1			Specimen 2			Specimen 3			Specimen 4			Specimen 5			
Name	**Line No.**	**Range & Type**	**Mean**	**SD**	**Range & Type**	**Mean**	**SD**	**Range & Type**	**Mean**	**SD**	**Range & Type**	**Mean**	**SD**	**Range & Type**	**Mean**	**SD**	**No. of Labs**
Ammonia, Serum																	
Initial Grouping by Reagent and Instrument																	
Siemens Dimension & Siemens Dimension EXL	1	257 - 314	P 285.5	12.2	17 - 37	C 26.6	4.1	-			-			-			33
Initial Grouping by Reagent																	
Ortho Vitros	2	233 - 284	P 258.5	15.7	0 - 20	C 9.5	0.8	-			-			-			10
Siemens Dimension	3	258 - 315	P 286.2	12.3	17 - 37	C 26.9	4.3	-			-			-			40
Initial Grouping by Sensitivity or Principle																	
All other methods	4	255 - 312	P 283.6	18.6	16 - 36	C 26.3	4.1	-			-			-			63
Dry Slide or Cartridge	5	233 - 284	P 258.5	15.7	0 - 20	C 9.5	0.8	-			-			-			10
Total Population																	
Whole Population	6	252 - 308	P 280.0	20.1	14 - 34	C 23.9	7.0	-			-			-			74